INCH-POUND

A-A-54699
13 July 1992

COMMERCIAL ITEM DESCRIPTION

THERMOMETER, SELF-INDICATING, BIMETALLIC, DIAL, 8-1/2 INCH

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-T-36349A which is cancelled.

This Commercial Item Description covers one type of self-indicating, dial-type thermometer, suitable for use in film-developing tanks.

Salient characteristics:

General. The thermometer shall be a bimetallic, dial, self-indicating type consisting of a stem and head with a metal clip on the stem for holding the thermometer in upright position against a film developing tank. The design shall provide means for recalibration without disassembling the head. The head shall be waterproofed. No visible liquid leakage shall be inside when immersed in 20 inches of water within 1/2 hour.

Material. The stem and head shall be made of corrosion-resisting steel AISI 300 series. As an alternate, nickel plated steel for the head is acceptable, but the stem shall be corrosion-resisting steel. The dial crystal shall be either in glass or plastic.

Style, design and dimensions. The head shall contain the dial face, 2 inches $\pm$ 1/4 in diameter, with stem 6 to 8 inches long perpendicular to the head. Range of scale shall be from at least $25^{\circ}F$ to $125^{\circ}F$ and shall cover an arc of not less than 290°. The graduation shall be at $1^{\circ}F$ intervals, each 5th mark shall be longer than the intervening marks. Each 10° shall be marked with Arabic numerals.

Accuracy. Shall be accurate to $\pm 1^{\circ}F$ from $60^{\circ}F$ to $85^{\circ}F$ and within $2^{\circ}F$ at scale extremes.

AMSC N/A FSC 6685

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Protective case. A protective case shall be provided. It shall be a spiral-wound fiberboard can with metal screw cap closure. As an alternate, a telescope type tube closure is acceptable. Suitable cushioning shall be provided to prevent movement of the thermometer within the case.

Identification marking. The thermometer shall be legibly and permanently marked with the name or registered trademark of the manufacturer.

Workmanship. The thermometer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One thermometer, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination and Testing. Examination and testing shall be conducted on the thermometer in accordance with commercial practice. The sampling for examination and testing shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6685-00-616-7950	THERMOMETER, SELF-INDICATING, BIMETALLIC, DIAL, 8-1/2 INCH.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6685-0880

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revisio letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification o requirements on current contracts. Comments submitted on this form do not constitute or imply authorization t waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54699	13 JULY 1992

3. DOCUMENT TITLE

THERMOMETER, SELF-INDICATING, BIMETALLIC, DIAL, 8-1/2 INCH

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. [illegible]	b. ORGANIZATION	
c. ADDRESS [illegible]	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (if applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center	(1) Commercial (215) 737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) ATTN: MSE 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete. U.S. Government Printing Office: 1990-704-119/20055 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:37Z
Check the source to verify that this is the current version before use.