INCH-POUND

A-A-54755
30 September 1992

COMMERCIAL ITEM DESCRIPTION

BANDAGE, GAUZE, SIX-PLY, 4-1/2 Inches Wide by 4 Yards

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers six-ply bandages suitable for use as a burn or surgical dressing.

Salient Characteristics:

Shall be sterile, roller-type elastic (stretchable) gauze bandages suitable for use as a burn or surgical dressing. Bandages shall be six-ply. Bandages shall be sterile when tested in accordance with the USP.

The gauze shall undergo treatment causing it to shrink and to have excess crimp in the warp direction. Excess crimp in filling direction may also be imparted to the gauze. The minimum thread count of the treated gauze shall be 13 warp by 6 filling threads per inch measured in the relaxed condition when tested as specified in method 5050 of Federal Test Method Standard 191. The finished treated gauze shall meet the requirements of Absorbent Gauze, USP, except for the gauze thread count and gauze weight. When cotton is not blended with rayon, the water extract shall not exceed 28 mg. A maximum of 0.06 percent polyester fiber may be used. The treated gauze shall be stretchable.

No adhesive, rubber nor sizing materials shall be used in the manufacture of the bandage. The bandage shall be made by folding the gauze to six-ply. The width of all plies shall be the same. Each bandage shall be rolled with the edges evenly aligned and the gauze shall not be stretched in the roll, but shall be wound with sufficient tension to keep the roll from gapping. The bandages shall be of one continuous piece and shall not be cut on the bias. The weight of a roll shall be a minimum of 29 grams. There shall be no exposed raw edges.

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

The width of the finished bandaged shall be 4-1/2 inches plus or minus 1/4 inch and the length shall be not less than 100 inches when measured within five minutes after unrolling. The stretched length of the finished bandage shall average no less than 147 inches with no individual length less than 137 inches when the ends of the bandage are clamped and the bandage stretched to its breaking point.

The finished bandage (six-ply) shall have a minimum breaking strength of 25 pounds in the warp direction when tested in accordance with method 5100 of Federal Test Method Standard 191 with the distance between the jaws of the tester 1-1/2 inches apart at the start of the test.

Bandages shall be clean. Bandages shall be free of loose threads, ravelings, lint cuts and tears.

Workmanship. The bandages shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

For Size 1, each rolled bandage shall be supplied in a sealed, peel-open package capable of maintaining sterility of contents unless package is damaged or opened. One package containing 100 individually packaged bandages, as specified, constitutes one unit.

For Size 2, each rolled bandage shall be supplied in a sealed, peel-open package capable of maintaining sterility of contents unless package is damaged or opened. One package containing 12 individually packaged bandages, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-058-3047	BANDAGE, GAUZE, SIX-PLY, 4-1/2 inches by 4 yard, 100s	1
6510-00-582-7992	BANDAGE, GAUZE, SIX-PLY, 4-1/2 inches by 4 yards, 12s	2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

PGC 06833

Project No. 6510-0846

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54755	30 SEP 1992

3. DOCUMENT TITLE

BANDAGE, GAUZE, SIX-PLY, 4½ Inches wide by 4 Yards

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete.* '98

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:38Z
Check the source to verify that this is the current version before use.