INCH-POUND

A-A-54768
6 November 1992

COMMERCIAL ITEM DESCRIPTION

LABEL, Add-On

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a blood group add-on label for use on blood bags.

Salient characteristics:

Shall be an add-on label for use with labels preapplied to blood bags. Print type, print size, color, ink, adhesive, paper and all other features and material shall conform to the latest edition of "Guidelines for the Uniform Labeling of Blood and Blood Components". Maximum label width shall be 1-3/4 inches and maximum length shall be 2-3/8 inches. The label shall contain information concerning only blood type, RH factor and applicable bar code. Color is to be black and white only. No clauses shall appear on label.

Workmanship. The label shall be free from defects which detract from its appearance or impair its serviceability.

Unit. ROLL (RO). For NSN's 7690-01-350-0350, 7690-01-350-5957 and 7690-01-350-5958; one roll containing two hundred and fifty labels, as specified, constitutes one unit.

For NSN's 7690-01-350-5959, 7690-01-237-1935 and 7690-01-240-5756; one roll containing one hundred labels, as specified, constitutes one unit.

For NSN's 7690-01-237-6069 and 7690-01-240-5753; one roll containing five hundred labels, as specified, constitutes one unit.

AMSC N/A FSC 7690

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one roll of labels as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one roll of labels. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one roll of labels. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL of 2.5, and sampling for tests shall be conducted at inspection level S-1, AQL 1.0.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS
7690-01-350-0350	LABEL, Add-on, "BLOOD GROUP B, RH POSITIVE" Legend 2-3/4 inches long by 1-3/4 inches wide, 250s
7690-01-240-5753	LABEL, Add-on, "BLOOD GROUP A, RH POSITIVE" Legend, 2-3/8 inches (6 cm) long, 1-3/4 inches (4.4 cm) wide, 500s
7690-01-350-5957	LABEL, Add-on, "BLOOD GROUP A, RH NEGATIVE" Legend, 2-3/8 by 1-3/4 inches, 250s
7690-01-350-5959	LABEL, Add-on, "BLOOD GROUP B, RH NEGATIVE" Legend, 2-3/8 inches by 1-3/4 inches, 100s
7690-01-237-6069	LABEL, Add-on, "BLOOD GROUP 0, RH POSITIVE" Legend, 2-3/4 inches (6 cm) long, 1-3/4 inches (4.4 cm) wide, 500s
7690-01-237-1935	LABEL, Add-on, "BLOOD GROUP AB, RH POSITIVE" Legend, 2-3/8 inches (6 cm) long, 1-3/4 inches (4.4 cm) wide, 100s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS
7690-01-240-5756	LABEL, Add-on, "BLOOD GROUP AB, RH NEGATIVE" Legend, 100s
7690-01-350-5958	LABEL, Add-on, "BLOOD GROUP O, RH NEGATIVE" Legend, 2-3/8 inches (6 cm) long, 1-3/4 inches (4.4 cm) wide, 250s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 7690-0091
Location: c:\word\nish\labelcid

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54768	06 November 1992

3. DOCUMENT TITLE

LABEL, Add-On

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial 215-737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:38Z
Check the source to verify that this is the current version before use.