INCH-POUND

A-A-54844
12 April 1993

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, BLOOD COLLECTING

The General Services Administration has authorized the use of this Commercial Item Description as a replacment for Military Specification MIL-N-42013A, which is cancelled.

This Commercial Item Description covers a sterile, disposable, double pointed hypodermic blood-collecting needle capable of multiple sampling from one injection site.

Salient characteristics:

Material. The needle cannula shall be fabricated from Class 3, Type 304 stainless steel in accordance with ASTM F899. The hub shall be fabricated from rigid polypropylene that meets the United States Pharmacopeia (USP) requirements for Class IV plastics. The shut-off sleeve of the needle shall be fabricated from latex rubber or natural rubber which meets the USP requirements of Elastomeric Closures for Injections. The protective sheaths shall be fabricated from rigid plastic.

Style and design. Shall consist of a thin wall 21 gauge double pointed needle with protective sheath on each end. The needle shall have a threaded hub designed to attach to a tube holder. The tube end of the needle shall be open at the proximal end and shall be covered by a rubber shut-off sleeve which shall be securely attached to the hub. The shut-off sleeve shall operate automatically as the blood collecting tube is forced against the needle displacing the shut-off sleeve. Withdrawal of the blood collecting tube shall result in resealing the blood collecting needle. The needle shall be 21 gauge by 1-1/2 ± 1/16 inches long on the patient end with the following dimensions (inches):

Overall length	Cannula ID/OD	Point length
2 9/16 ± 1/8	0.022 ± 0.002/0.032 ± 0.003	0.151 ± 0.012

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The hub shall have the following dimensions (inches):

Hub Diameter	0.275 ± 0.005
Hub Thread, Double	0.185 - 18 NS
Major Diameter	0.185 ± 0.002
Minor Diameter	0.139 + 0 - 0.004
Pitch Diameter	0.164 + 0 - 0.004

Non-toxic and pyrogen-free. The needle shall be non-toxic and pyrogen free when tested in accordance with USP.

Sterility. The needle shall be tested for sterility in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.0001.

Resistance to separation. The needle cannula shall not separate from the hub when subjected to a 5 pound axial tensile force exerted for one minute (See Note 2).

Workmanship. The needle shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one thousand individually packaged sterile hypodermic needle together with ten blood collecting tube holders, as specified, constitutes one unit. Cannula used to enter the vein shall be protected by a sheath in such a manner that by turning the sheath the hub shall be screwed into the orifice of the tub, holder without the use of instruments. The individual package shall be a suitable container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one blood collecting hypodermic needle and one holder as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one blood collecting hypodermic needle and one holder. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Numbers and their respective Item Identifications is covered by this document:

National Stock Number	Item Identification
6515-01-008-7103	NEEDLE, HYPODERMIC, BLOOD COLLECTING, Multiple Sample, Double Pointed, 21 gauge, 1000s

NOTE 2: As alternate test, such as a pull-to-destruction test is acceptable in lieu of the 5 pound separation test, provided product found acceptable of the pull-to-destruction test also passes the 5 pound separation test.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5424
Location: C:\MARK\BLOOD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54844	2 DOCUMENT DATE (YYMMDD) 12 April 1993

3 DOCUMENT TITLE
NEEDLE, HYPODERMIC, BLOOD COLLECTING

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)*

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial (215)7372870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.