METRIC

A-A-54846
1 April 1993

COMMERCIAL ITEM DESCRIPTION

INDICATOR, OXYGEN, ANAEROBIC CULTURE APPARATUS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable anaerobic indicator.

Salient Characteristics:

Shall be a disposable anaerobic indicator for a self-contained anaerobic apparatus or other type with a transparent jar, used for cultivation of anaerobic micro-organisms.

Shall be a chemical saturated cotton pad that is supplied in a sealed peelable plastic-aluminum foil laminate overwrap. The cotton pad shall be saturated with a methyl blue solution having characteristics of exhibiting a blue coloration when exposed to air or oxygen, and changing to a colorless condition in an anaerobic atmosphere devoid of oxygen. The colorless reaction shall be reversible by returning to the blue coloration as a result of re-exposure to air or oxygen. The nominal dimensions of the cotton pad shall be 1.6 cm wide x 5.08 cm long. The nominal dimensions of the plastic-aluminum overwrap holding the cotton pad shall be 3.49 cm wide x 11.43 cm long.

Shall be supplied so that the cotton pad is made ready for use by peeling apart the foil covering at one end to halfway down to expose the saturated cotton pad. The exposed cotton pad in the open foil covering is then inserted into an anaerobic jar near a side that is transparent and suitable for external viewing.

AMSC\NA FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Shall be suitable to indicate when anaerobiosis is achieved within a properly prepared anaerobic jar, as a result of the anaerobic atmosphere internally developed by the generation of hydrogen and carbon dioxide gases. Shall be suitable for use in anaerobic jars having nominal dimensions of 22.9 cm overall height x 13 cm outside diameter at its midpoint.

Workmanship. The indicator and overwrap shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 100 indicators, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted in accordance with commercial practice on the indicator. Sampling for inspections and testing shall be in accordance with commercial practice.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Ordering data. Expiration dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6640-00-145-1150	INDICATOR, OXYGEN, ANAEROBIC CULTURE APPARATUS, Disposable, 100s

This indicator is intended for use with the following stock listed components forming an anaerobic apparatus:

NSN	ITEM IDENTIFICATION
6640-00-001-6670	LID, ANAEROBIC CULTURE APPARATUS
6640-00-926-7655	JAR, ANAEROBIC CULTURE APPARATUS
6640-00-935-6863	GENERATOR, CARBON DIOXIDE-HYDROGEN, ANAEROBIC CULTURE APPARATUS, Disposable, 100s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1804

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54646	1 APRIL 1993

3 DOCUMENT TITLE
INDICATOR, OXYGEN, ANAEROBIC CULTURE APPARATUS

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: MSE	(1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.