INCH-POUND

A-A-54851
16 April 1993

COMMERCIAL ITEM DESCRIPTION

LITTER-SPLINT, EXTRICATION

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a radiolucent, full-body splint.

Salient Characteristics:

The extrication litter-splint shall be a radiolucent, full-body splint used for total immobilization, as well as extrication and transport of injury victims. The splint shall be furnished in the following two colors:

* Olive drab color.

* Yellow color.

The splint shall be constructed of a high density polyethylene outer shell with an injected foam core of rigid polyurethane foam. The polyethelene shell shall have a minimum density of 0.930 grams per cubic centimeter when tested in accordance with ASTM D1505. The polyethylene shell shall not splinter or crack, shall be resistant to chemicals, gasoline, blood and other body fluids, with no discoloration or damage, shall be fungus resistant, and shall be easily cleaned with soap and water.

The splint dimensions shall be approximately 66 inches long by 14 inches wide by 3-1/2 inches thick. The splint shall have a minimum of 6 raised handholds, a minimum of three on each side. The handholds shall be raised to allow extra space for those litter bearers wearing gloves. The splint shall be equipped with or without split legs and shall fit within a stokes wire basket (without split leg design) for

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

storage, or while transporting a patient. The splint shall weigh no more than 18 pounds with a load capacity of 300 pounds.

The splint shall have three webbing restraint straps approximately 2 inches wide, made of commercial grade washable nylon with a breaking strength of 5000 pounds minimum. The straps shall be located at a minimum of three different horizontal positions of the body. The straps shall use fasteners made from a washable 2 inches wide hook and pile (velcro or equal) material. The locations and lengths of the restraint straps and fastener material shall be suitable for immobilizing either a small child or large adult (35 to 300 pounds) with adjustment being accomplished by plated steel or sufficiently tough plastic slides.

Slide position adjustment and fastener placement shall be adequate to maintain immobilization of any size patient while the patient is undergoing x-ray examination without the slide or fastener image interfering with the patient image. New adjustment may be allowed to facilitate different views of the patient. The radio translucent splint, complete with all straps, fasteners, slides, or pins, shall not interfere with CT scans or magnetic resonators.

The splint shall be equipped with a removeable padded head restraint of sufficient size and adjustment capability for use with either a combat helmeted or unhelmeted patient.

Workmanship. The extrication litter-splint shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One litter-splint of the color specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one extrication litter-splint.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one extrication litter-splint.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document.

6530-01-193-5037 (olive drab color)
6530-01-199-1969 (yellow color)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6530-2297

Location: SPLTC13B.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54851	2. DOCUMENT DATE (YYMMDD): 16 April 1993

3. DOCUMENT TITLE: LITTER-SPLINT, EXTRICATION

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING activity

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. 98790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.