INCH-POUND

A-A-54862
30 April 1993

COMMERCIAL ITEM DESCRIPTION

TUNING FORK SET

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-T-36184A which is cancelled.

This Commercial Item Description covers a tuning fork set suitable for use as a diagnostic aid during medical and neuropsychiatric examinations.

Salient characteristics:

General. The tuning fork set shall consist of one each of the following: tuning fork C128, tuning fork C256, tuning fork C512, tuning fork C1024 and tuning fork C2048.

Material. Material shall be as follows:

Body and fork proper - 2024 T3 or T4

Weight on arms - 2107 T4 or T6 or 2011 T3 aluminum.

As an alternate, the weight may be the same material as the body and fork proper.

Style, design and dimensions. Tuning fork set shall be in accordance with Figure 1, except dimensions shown shall be considered "reference" dimensions for outline purposes only.

Handles. Handles shall be an integral part of the forks and of the same diameter. They shall be slightly rounded and polished to a smooth surface or finished to a satin smooth finish.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Prongs. The prongs of the C128 and C256 tuning forks shall be fitted with cylindrical weights fixed with screws, and shall be non-adjustable.

Vibration. Each fork shall show a maximum intensity of response and a maximum period of sustained audible vibration for a given energy applied to excite the fork, and shall give the same frequency for both air and bone conduction tests as indicated in the table below. When struck sufficiently hard on a non-metallic object, the minimum period of sustained vibration shall be:

Tuning Fork	Air Conduction	Bone Conduction
C-128	37 seconds	22 seconds
C-256	17 seconds	17 seconds
C-512	35 seconds	17 seconds
C-1024	30 seconds	----------
C-2048	83 seconds	----------

Frequency, tone and tolerance. When determined with a suitably calibrated frequency counter, the frequency, tone and tolerance shall be as follows:

Tuning fork	Tone	Frequency (cycles per second)	Tolerance (cycles per second)
C-128	C	128	6
C-256	C	256	12
C-512	C	512	15
C-1024	C	1024	20
C-2048	C	2048	40

Case. Shall be made of a material able to withstand high impact such as wood, steel, or vinyl. Interior shall be arranged to hold each fork of the set in a definite position. Case shall be equipped with clasp to insure closure.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Stamping. Each tuning fork shall be legibly and permanently stamped with the letter "C" and the specific number of cycles per second, together with the name or trademark of the manufacturer.

Workmanship. The tuning fork sets shall be free from defects which detract from their appearance or impair their serviceability. The tuning forks shall be symmetrical in shape and shall have smoothly finished surfaces.

Unit. Set (SE). One set consisting of five (5) tuning forks along with carrying case, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one tuning fork set. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be tuning fork set. Sampling shall be Inspection level S-2 with an AQL of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor Certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Number is covered by this document:

NSN	Item Identification
6515-00-338-9200	TUNING FORK SET

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5336

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.




FIGURE 1 TUNING FORK SET

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54862	30 April 1993

3. DOCUMENT TITLE

TUNING FORK SET

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON
c. ADDRESS (Include Zip Code)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.