METRIC

A-A-54865
30 June 1993

COMMERCIAL ITEM DESCRIPTION

SYRINGE, IRRIGATING, BLADDER EVACUATING
Toomey

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a glass Toomey bladder evacuating irrigating syringe.

Salient characteristics:

Material. The syringe plunger and barrel shall be fabricated from borosilicate glass. The syringe tip adapter shall be fabricated from corrosion-resistant metal.

Style and design. The syringe shall be a 50 cubic centimeters (cc) Toomey type evacuating syringe with syringe tip adapter. The adapter shall be designed to fit all standard cystoscopes and resectoscopes. The syringe and adapter shall be capable of withstanding repeated autoclaving without corrosion or other detrimental effects.

The barrels, plungers and adapters shall be interchangeable between syringes without any leakage during use. The inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks or other imperfections. The plunger shall glide through the barrel while maintaining a fluid-tight fit. The barrel shall be transparent.

The outside surface of the barrel shall be marked with main graduations every 5 cc and subgraduations every 1 cc. Main graduation markings shall be at least 33 percent longer than subgraduations. Each 10cc graduation shall also have the numerical volume indicated. The barrel shall have the name or registered trademark of the contractor. Each syringe shall be packaged in an individual container that shall protect the syringe during shipment and storage.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Leakage. The adapter shall be attached to the syringe and then attached in a horizontal position to a water pressure source having a pressure gauge. The water pressure shall be brought up to 25 psig. The plunger shall be withdrawn to the 50 percent point of the graduated capacity where it shall remain for 10 seconds. The plunger shall then be withdrawn to the 100 percent point and held for another 10 seconds. During each 10 second period there shall not be enough water to pass the connection between the adapter and the syringe or between the plunger tip and the barrel of the syringe to produce a falling drop. After each successive test any water that may have passed shall be wiped with a dry cloth and the next test applied. The syringe shall be removed from the pressure source and the water shall be ejected at the rate of 1/2 cc per second until the plunger reaches the zero mark. The adapter shall not leak or separate from the syringe.

Workmanship. The syringe shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One syringe as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one syringe as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one syringe. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

National Stock Number	Item Identification
6515-00-926-1487	SYRINGE, IRRIGATING, BLADDER EVACUATING, Toomey 50 cc

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5573
Location: c:\mark\syringe4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

RECOMMENDED CHANGE	1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
	A-A-54865	930531

3 DOCUMENT TITLE

SYRINGE, IRRIGATING, BLADDER EVACUATING, Toomey

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
c ADDRESS (Include Zip Code)		

7 PREPARING ACTIVITY

a NAME	b TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.