METRIC

A-A-54868
17 May 1993

COMMERCIAL ITEM DESCRIPTION

MICROSCOPE, OPTICAL

Stereoscopic

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a portable stereoscope optical zoom microscope with a carrying and storage type case, for biological examination and microdissection of small specimens.

Salient characteristics:

The microscope shall be a stereoscopic optical zoom microscope with substage illumination, for biological examination and microdissection of small specimens, such as insects, that are opaque, transparent, and semi-transparent.

A rigid carrying and storage case with suitable brackets for securely holding the microscope assembly and accessory components shall be supplied with the microscope. The case shall have a hinged cover and shall include latches.

Optics shall be provided that include a 0.7X through 3X continuous variable zoom magnification range, interchangeable pairs of 15X and 20X widefield eyepieces, and a 2X supplementary (auxiliary) lens.

The microscope shall be suitable for biological examination and dissection using either top mounted external oblique illumination or substage illumination (Note that only the substage illuminator shall be supplied with the microscope).

AMSC N/A FSC 6650

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The stereoscope optical microscope shall include the following:

* A microscope body with internal zoom optics, 30 degree inclined binocular tubes, and a knob (dial) control for variable zoom magnification.

* A balanced support stand consisting of a horizontal base, an upright stand with focusing controls, and a ring support arm for holding the microscope body.

* A substage illuminator base with an adjustable reflecting mirror and illumination system.

* A pair of 15X wide field eyepieces with eyeguards. At least one of the eyepieces shall have an acuity equalizer.

* A pair of 20X wide field eyepieces with eyeguards. At least one of the eyepieces shall have an acuity equalizer.

* A 2X supplementary (auxiliary) lens for attachment to the microscope body.

* A clear glass stage plate.

* A set of opaque black and white stage plates, or as an alternate, a single stage plate with opaque black and white faces on reversed sides.

* Two (2) large size stage holding clips that are secured to the base of the stand by thumbscrews.

* A protective dust cover for the assembled microscope and support stand.

* The unit shall be supplied with a microscope body with inclined binocular tubes. The binocular tubes shall be inclined 30 degrees from the vertical and shall include an interpupillary adjustment. The microscope body optics shall have a variable continuous magnification range that includes 0.7X to 3X. The microscope body shall include a graduated knob (dial) control for adjusting the magnification. The microscope body shall be supported in the ring support arm of the microscope stand. When properly seated in the ring support arm, the microscope body shall be 360 degree positionable and also securable in any desired position by tightening a locking screw that shall be included with the ring support arm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The microscope shall be supplied with a balanced support stand. The base of the support stand shall include a suitable cutout for insertion of interchangeable stage plates. The base shall also include threaded holes for attaching the two stage clips. The vertical upright support shall include a knob (dial) controlled focusing system, which includes a rack and pinion mechanism, for adjusting the working distance of the microscope through the height adjustment of the support ring. The motion tension on the knob (dial) control shall be adjustable in accordance with instructions that shall be included in the service data that is supplied with the microscope. The base shall include suitable means for attaching a substage illuminator housing.

The microscope shall be supplied with a substage illuminator housing that can be secured to the bottom of the base of the support stand. The substage illuminator shall include a reflecting mirror that is tilted by turning a knob (dial) control that is attached on the outside of the housing. The substage illuminator housing shall include a suitable fitting for inserting an illuminator for use in conjunction with the mirror.

The microscope shall be supplied with a tungsten halogen 6 volt, 10 or 20 watt illumination system that can be set in the substage illuminator housing for transmitting a beam of light to the reflecting mirror. The illumination system shall be operable from a 115 volt, 50/60 Hz AC supply source, and shall include a three-wire power cord that terminates with a three-blade male grounding plug.

Service data. Two copies each of one or more service data booklets shall be supplied with the microscope. The service data shall contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit, other than common fasteners shall be identified in the service data and part listing by the manufacturer's part number.

Workmanship. The microscope, with the stand, optical components, and carrying case shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One stereoscopic microscope, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the microscope, stand, optical components and carrying case in accordance with commercial practice. In addition each microscope and focusing control shall be checked to demonstrate their capability for proper operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

6650-00-973-6945

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6650-0192

Location: MICRO14C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54868	2. DOCUMENT DATE (YYMMDD): 17 May 1993

3. DOCUMENT TITLE: MICROSCOPE, OPTICAL Stereoscopic

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 98/TPC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.