INCH-POUND

A-A-54870
17 May 1993

COMMERCIAL ITEM DESCRIPTION

CAPS, OPERATING, SURGICAL

(Woman's)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-C-37268 which is cancelled.

This Commercial Item Description covers a woman's disposable surgical operating cap.

Salient characteristics:

The surgical operating cap shall be fabricated from nonwoven fabric consisting of 100 percent rayon or 100 percent spunbonded polypropylene, constructed circular in shape, and furnished in the following two sizes:

Size 1 - 12 inches in overall diameter, intended for wear by women with bouffant hair styles.

Size 2 - 10 inches in overall diameter, intended for wear by women without bouffant hair styles.

The rayon material shall have an acrylic latex binder and a fire retardant treatment. The material shall be of quality to withstand sterilization by steam without impairing the serviceability of the cap. Bacteriostatic agents, if nontoxic and harmless in the concentration employed, may be added. The finished nonwoven fabric shall be non-irritating and non-toxic when inspected by any standard method. The fabric shall be non-pilling, free of lint and objectionable odor when wet or dry, and noiseless. The cap fabric shall conform to the following requirements:

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Rayon

Weight, grams per square yard	
Minimum	13
Maximum	22
Tensile strength, pounds minimum	
Machine direction	3.4
Cross direction	0.35

Polypropylene

Weight, grams per square yard, minimum	12.5
Tensile strength, pounds minimum	
Machine direction	2.0
Cross direction	2.0
Bursting strength, pounds minimum	14.0

Rayon & Polypropylene Materials

Flame resistance, seconds minimum	3.0

Colors shall be light shades of blue, green, or grey, and may include a pattern of small white stripes or flecks.

Flame resistance shall be determined using National Fire Protection Association (NFPA) Inspection Method 702, Flammability of Wearing Apparel.

The cap shall be of one piece construction, circular in shape, with an elastic band around the circumference to hold the cap on the wearer's head. The elastic band shall be one round, white strand of extruded rubber covered with white cotton yarns. The elastic band shall have a minimum elongation of 210 percent before breaking.

The size of the fabric for the size 1 cap shall be a minimum of 24 inches in diameter. The edge to which the elastic is to be attached shall be folded approximately 1/2 inch. The elastic band shall be attached with overedge stitching and the shirrings or gatherings shall be uniform. As an alternate, the size of the fabric shall be 19 plus or minus 1/2 inches wide by 21 plus or minus 1/2 inches long. The elastic band shall be attached to each 21-inch edge with

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

overedge stitching. Each 19-inch edge shall be gathered uniformly and securely clinched, 3/8 inch plus or minus 1/8 inch from the end, by a centered metal fastener, or by tacking with cotton thread. The metal fastener, if used, shall be fabricated from non-corrosive, anti-spark metal. The relaxed head opening of the cap shall be 6-1/2 inches plus or minus 1/2 inch. The overall diameter shall be a minimum of 12 inches.

The size of the fabric for the size 2 cap shall be a minimum of 19 inches in diameter. The elastic band shall be attached to the periphery as specified above for the size 1 cap. As an alternate, the size for the fabric shall be 19 inches plus or minus 1/2 inch square. The elastic band shall be attached to two opposite sides. The two remaining sides shall be gathered uniformly and secured as specified above for the size 1 cap. The relaxed head opening shall be 6-1/2 inches plus or minus 1/2 inch. The overall diameter shall be a minimum of 10 inches.

The overedge stitching used in the cap shall be stitch type 503, 504, 505, or a combination of 101 and 102. The bight of the overedge stitching shall be 1/8 to 1/4 inch from the edge of the fabric. All regular sewing shall be performed with stitch type 301. All sewing shall be accomplished with a minimum of 6 stitches per inch. The seam allowance may be on the inner or outerside of the cap. Each thread break, skip, and run-off shall be overstitched not less than 1/2 inch. Skipped stitches, or thread breaks, may be repaired using stitch type 301. The thread tension shall be maintained so that there shall not be any loose or tight stitching and the lock shall be embedded in the materials sewn together. All thread ends shall be trimmed to a maximum length of one inch.

The cap shall not contain any spot, stain, hole, abraded area, tear, ravelings, cut, mend, scorch, or needle chew. The fasteners, if used, shall be clinched without distortion, damage, splitting, or cutting of the fabric. The fasteners shall not be malformed, misaligned, fractured, broken, or distorted, nor contain any corrosion, scale, pit, dent, crack, burr, sliver, or sharp edge.

Workmanship. The surgical operating cap shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing five hundred caps of size 1 or one package containing one thousand caps of size 2, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 4.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical operating cap.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one surgical operating cap.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

6532-00-250-5041 (Size 1)
6532-00-250-5042 (Size 2)

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6532-1206

Location: CAPS13C.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54870	2. DOCUMENT DATE (YYMMDD): 17 May 1993

3. DOCUMENT TITLE: CAPS, OPERATING, SURGICAL (Woman's)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870	(2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 96790

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.