INCH-POUND

A-A-54872
1 June 1993

COMMERCIAL ITEM DESCRIPTION

BLOOD EXCHANGE-TRANSFUSION KIT, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a ready to use disposable blood exchange-transfusion kit suitable for use in the treatment of Rh blood incompatibility and for transfusion of blood without cross infection.

Salient characteristics:

General. The blood exchange-transfusion kit shall be furnished in a plastic tray approximately 15-1/2 inches long, 9-1/8 inches wide and 1-1/8 inches in height. The blood exchange-transfusion kit shall contain the following components as a minimum:

Quantity	Component
1	Blood Transfusion Set, with Nylon Blood Filter, Tubing and Control Clamp
1	Ampul of Calcium Gluconate Injection, USP, 10 percent, 10 ml
1	Container for Waste Blood with Drawstring Closure, 2000 ml, and with 100 ml Calibrations
1	Exchange-Transfusion Record
1	Extension Tube, 33 inches long
1	Stopcock Assembly (Stopcock with 40 inch extension tube)
3	Gauze Sponges, 2 by 2 inches, 12 ply

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:39Z
Check the source to verify that this is the current version before use.

Quantity	Component
1	Ruler, Plastic, 15 cm
1	Polypropylene Syringe, 5 ml with 23 gauge by 1 inch needle
2	Polypropylene Syringes, Luer-Lock, 20 ml
1	Umbilical Catheter, 5 Fr, 14 inches long with 0.22 ml capacity
1	Umbilical Catheter, 8 Fr, 14 inches long with 0.88 ml capacity

Blood transfusion set. The blood transfusion set shall consist of a piercing device, filter, housing, nylon filter, 60 inches of plastic tubing, 2-1/2 inches of natural rubber tubing on proximal end (next to filter), 4 inches of rubber tubing on distal end (next to needle adapter), needle adapter and a flexible aluminum control clamp. Filter mesh area shall not be less than 4 square inches with a mesh size of 70 to 100 mesh sieve size United States Pharmacopeia (USP).

Pyrogens, safety and biological tests. All rubber and plastic components shall be sterilizable, free from pyrogens in accordance with the USP, safe in accordance with the USP (nontoxic) and shall contain not more than 5 parts per million (ppm) of lead, barium and cadmium. The plastic shall meet the requirements of the USP for Class IV plastics.

Material. Plastic tubing shall be transparent, colorless polyvinyl. The piercing device and needle adapter shall be fabricated from suitable rigid plastic. Needles shall be fabricated of Type 304 corrosion-resisting steel in accordance with ASTM Designation F899.

Sterility. Components and inside surfaces of the tray shall be sterile in accordance with the USP or AAMI guidelines. Sterility Assurance Level shall be 0.0001.

Instructions. Each kit shall be furnished with a list of the kit contents and procedural instructions for use.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:39Z
Check the source to verify that this is the current version before use.

Pressure tests. The kit shall not leak nor shall any damage occur when subjected to an internal hydrostatic pressure of 8 psig for 30 seconds.

Workmanship. The kit shall be free from defects which detract from its appearance or impair its serviceability. Components and tray shall be free from burrs, flash, cracks, corrosion or foreign material.

Unit. Package (PG). One package containing five individually packaged sterile blood exchange-transfusion kits as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one blood exchange-transfusion kit as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one blood exchange-transfusion kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.

regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment fror receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-369-5060	BLOOD, EXCHANGE-TRANSFUSION KIT, Disposable

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5587
Location: c:\word\doc\nish\bloodexc

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:39Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	2-A-54872	1 June 1993

3. DOCUMENT TITLE
BLOOD EXCHANGE TRANSFUSION KIT DISPOSABLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN DPSC-MSE	(1) Commercial (215)7372870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198.29C

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:39Z
Check the source to verify that this is the current version before use.