METRIC

A-A-54873
21 May 1993

COMMERCIAL ITEM DESCRIPTION

BOXES, MICROSCOPE SLIDE, PLASTIC

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification NNN-B-585B, which is cancelled.

This Commercial Item Description covers two sizes of plastic microscope slide boxes suitable for accommodating 25 and 100 glass slides respectively.

Salient characteristics:

Classification. The microscope slide boxes shall be of the following sizes as specified under notes:

Size 1 - Large, 100 slides
Size 2 - Small, 25 slides

Material. Size 1 and Size 2 boxes shall be fabricated of phenolformaldehyde, polypropylene or polystyrene plastic and shall be brown, white, black or gray in color. Cover and body shall be the same color and shall be fabricated of the same material.

Design and Construction.

Size 1. The box shall consist of a body and hinged cover which shall provide a flush fit on the top edge of the body wall. Phenolformaldehyde and polystyrene boxes shall have two hinges on one long side and a slide clasp or other suitable locking feature on the opposite long side to assure convenience and ease in opening and closing. Polypropylene boxes shall have one 15 cm length (minimum) hinge in center of one long side and suitable clasp or other locking feature on the opposite long side. The hinges for phenolformaldehyde and polystyrene boxes shall be formed

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

by metal pins driven into plastic hinge leaves which shall be integral with the cover and body or shall be secured to the cover and body by metal pins. The hinges and locking devices for polypropylene boxes shall be integral parts of the body and cover to form a one-piece box. Pins and other hardware shall be brass or steel (plated with nickel or chrome) or nickel.

Body. The interior of the body shall be fitted with three slide retainers molded integrally with the body. One slide retainer shall be along each of the two long sides and one shall be along the center long axis. The slide retainers shall have slots for 100 microscope slides in two rows of fifty slides each. The two retainers along the sides of the box shall be slotted vertically at equidistant intervals on the inner face of each retainer shall be slotted vertically on both opposing faces in lines with the corresponding slots of the slide retainers. Both center and side retainers shall be reinforced to prevent cracking and splitting. A soft cork or suitable foam plastic material shall be firmly affixed to the inside bottom to act as a cushion for reduction of breakage and rattling of the slides. Outside bottom surface shall make four-point contact with a plane surface to prevent tilting and jiggling. The box shall be capable of accommodating one hundred glass microscope slides 25 millimeters (mm) wide by 2 mm thick by 73.5 mm long as well as slides of the same width and thickness, but 77 mm in length. Internal height of the box shall be such that the slides shall project above the top edge of the uncovered body by at least 5 mm.

Size 2. The box shall consist of a body and a cover. The body and cover shall be so constructed that they form a flush fit when the box is closed. The cover shall be either removable or connected to the body by a hinge and shall have a suitable device to effect secure closure. The hinge of hinged box shall be 8.75 cm (minimum) long in center of one of the long sides of the box.

Body. The interior of the body shall have two slide retainers which are integrally molded to the long walls. Retainers shall have slots for twenty five microscope slides. The two retainers shall be slotted vertically at equidistant intervals and each slot shall be in line with the corresponding slot of the opposite retainer. Each pair of slots shall be capable of accommodating one slide 25 mm

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

wide by 2 mm thick by 73.5 mm long as well as slides of the same width and thickness and 77 mm in length. Internal height shall be such that the slides project over the top edge of the uncovered body by 6 ± 1 mm. The slides shall be retained in their proper slots, when the closed box is placed in any position. A cork or cork-like lining shall be permanently affixed to and completely cover, the entire interior bottom except for the slide retainers and slots. The exterior bottom surface shall make four-point contact with a plane surface to prevent tilting.

Dimensions. Dimensions shall be as shown in following table:

	Size 1	Size 2
Length	22.2cm ± 1.3cm	13.9cm ± 2.5cm
Width	16.5cm ± 1.3cm	6.75cm ± 0.6cm
Height (exclusive of four-point contact)	3.8cm ± 0.25cm 0.4cm	3.1cm ± 0.3cm
Wall thickness	1.5cm minimum	1.5cm minimum

Marking. Marking for slide identification.

Numbering and label, Size 1. A printed, numbered index label shall be firmly attached to the inside of the lid with corresponding numbers on the cork-type cushion on the inside bottom of the slide compartment to permit ready reference to the slides. The numbered index label shall cover the entire surface of the lid and shall consist of two rows of numbered lines from 1 through 50 in one row, and from 51 to 100 in the other row. Numbers designating each slide emplacement may be printed on top of the central divider in lieu of the numbering on the cork cushion. Numbers and lines shall be distinct and legible.

Numbering and label, Size 2. A printed, numbered index label shall be firmly attached to the inside of the lid with corresponding number on the cork lining on the bottom of the slide compartment to permit ready reference to the slides. The index label shall consist of one row of numbered lines, from 1 through 25,and shall cover the entire surface of the lid. Numbers and lines shall be distinct and legible. Numbers designating each slide emplacement may also be printed or molded adjacent to the upper edge of the slide partitions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Marking for identification. Each box shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Workmanship. The box shall be free from defects which detract from its appearance or impair its serviceability. The box shall be free from cracks, chips, breaks, flash, sprues, warping, sharp edges, distortion and foreign matter.

Unit. Each (EA). One microscope slide box of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one microscope slide box as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one microscope slide box. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, Item Identifications and Sizes are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6640-00-408-9915	Box, Microscope Slide, Plastic, 100 slide	1
6640-00-684-1345	Box, Microscope Slide, Plastic, 25 slide	2

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS
GSA-FSS-FCGC

Project Number: 6640-1822
Location: c:\word\doc\nish\boxescid

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54873	21 May 1993

DOCUMENT TITLE

BOEX, MICROSCOPE SLIDE, PLASTIC

NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)
Defense Personnel Support Center ATTN: DPSC-MSE	(1) Commercial: (215)7372870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041 3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — *Previous editions are obsolete* — 198-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.