INCH-POUND

A-A-54883
18 June 1993

COMMERCIAL ITEM DESCRIPTION

ARCH SUPPORT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-A-3327A, which is cancelled.

This Commercial Item Description covers arch supports furnished in pairs.

Salient characteristics:

Classification. The arch supports shall be of following types:

Type I - Combination (Longitudinal and Transverse)
Type II - Metatarsal (Transverse)
Type III - Plantar (Longitudinal)

Material. The arch cushions shall be fabricated from a open cell, cellular rubber Type I, Class A, Grade 4 in accordance with ASTM Designation 1056.

The stiffener shall be fabricated from a phenolic impregnated fabric.

The adhesive shall be fabricated of rubber mixed with pitch and resins dissolved in naphtha.

The top covering leather shall be mildew resistant case leather, full grain cattlehide or vegetable tanned cattlehide strap leather. The under leather shall be full grain cattlehide which is not treated for mildew resistance. The color be natural or brown.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Style and Design. The Type I combination style shall have a heel area which is slightly cupped. The longitudinal area shall be positioned between the talus and navicular on the medial side of the foot. The transverse area (metatarsal pad) shall be located 1/4 inch proximal to the Metatarsal heads (end of support). All patterns shall taper to a thin edge no greater than 3/32 inch at the distal and lateral borders.

The top covering leather shall be 3 ± 1 ounces thick. The under leather shall be 3/64 inch thick (minimum).

The Type II Metatarsal cushion shall have a heart shaped pad 3/16 ± 10 percent inch thick for all sizes and shall be located 1/4 inch proximal to the metatarsal heads (end of support).

The Type III plantar cushion shall have a slightly cupped heel area. The longitudinal area shall be positioned between the talus and navicular on the medial side of the foot. The following thickness in inches (± 10 percent shall apply):

Type	Size	Heel	Longitudinal	Transverse
I (Female)	9	5/32	1/2	3/16
I	6	1/8	3/8	1/8
I	7	1/8	7/16	1/8
I	8	5/32	1/2	1/8
I	9	5/32	15/32	1/8
I	10	5/32	15/32	3/16
I	11	5/32	5/8	3/16
I	12	3/16	11/16	3/16
III	6	1/8	3/8	----
III	7	1/8	7/16	----
III	8	5/32	1/2	----
III	9	5/32	15/32	----
III	10	5/32	15/32	----
III	11	5/32	5/8	----
III	12	3/16	11/16	----

Elongation. For case leather and vegetable tanned cattlehide strap leather at least 80 percent of the samples tested shall have an elongation less than 25 percent at a 25 pound load.

For full grain cattlehide at least 80 percent of the samples tested shall have an elongation of 25 to 60 percent at a 25 pound load. (See Note 2)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Colorfastness. The top covering leather shall be tested for colorfastness. The leather shall be tested for resistance to dry and wet cracking. Staining of the dry cloth shall be greater than or equal to Munsell value 8.5 and staining of the wet cloth shall be greater than or equal to Munsell value 6.5. (See Note 2)

Cracking. A sample approximately 6 inches by 1 inch representative of the leather used shall be subjected to a test for resistance to cracking on being bent by mechanical means. The presence of cracking shall be cause for rejection. (See Note 2)

Water soluble. Top covering leather fabricated from case leather or vegetable tanned cattlehide shall be tested for water soluble material. The leather shall have a maximum of 18 percent water soluble material.

Identification Marking. Each cushion shall be permanently marked with the size, type, male or female, right or left and the name or registered trademark of the contractor.

Workmanship. The arch support shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Pair (PR). One pair consisting of one left and one right arch support of the size and type specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one pair of arch supports as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one pair of arch supports. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

NOTE 1: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification	Type
6515-00-299-8004	ARCH SUPPORT, PLANTAR, MALE SIZE 8	III
6515-00-299-8005	ARCH SUPPORT, PLANTAR, MALE SIZE 11	III
6515-00-299-8009	ARCH SUPPORT, PLANTAR, MALE SIZE 12	III
6515-00-299-8012	ARCH SUPPORT, METATRSAL, MALE SIZE 10	II
6515-00-299-8026	ARCH SUPPORT, PLANTAR, MALE SIZE 9	III
6515-00-299-8027	ARCH SUPPORT, PLANTAR, MALE SIZE 10	III
6515-00-299-8028	ARCH SUPPORT, PLANTAR, MALE SIZE 6	III
6515-00-299-8029	ARCH SUPPORT, PLANTAR, MALE SIZE 7	III
6515-00-299-8040	ARCH SUPPORT, COMBINATION, FEMALE SIZE 9	I
6515-00-299-8060	ARCH SUPPORT, COMBINATION, MALE SIZE 6	I
6515-00-299-8061	ARCH SUPPORT, COMBINATION, MALE SIZE 7	I
6515-00-299-8062	ARCH SUPPORT, COMBINATION, MALE SIZE 10	I
6515-00-299-8063	ARCH SUPPORT, COMBINATION, MALE SIZE 11	I
6515-00-299-8064	ARCH SUPPORT, COMBINATION, MALE SIZE 8	I
6515-00-299-8065	ARCH SUPPORT, COMBINATION, MALE SIZE 9	I
6515-00-299-8066	ARCH SUPPORT, COMBINATION, MALE SIZE 12	I
6515-00-299-8071	ARCH SUPPORT, METATARSAL, MALE SIZE 11	II
6515-00-299-8072	ARCH SUPPORT, METATARSAL, MALE SIZE 8	II
6515-00-299-8073	ARCH SUPPORT, METATARSAL, MALE SIZE 9	II
6515-00-299-8075	ARCH SUPPORT, METATARSAL, MALE SIZE 7	II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

NOTE 2: Acceptable test for colorfastness, elongation cracking and water soluble are described in Federal Test Method Std. No. 311.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5461
Location: c:\mark\arch\dawne

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54883	2 DOCUMENT DATE (YYMMDD) 18 June 1993

3 DOCUMENT TITLE
ARCH SUPPORT

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial: (215)7372870 (2) AUTOVON: 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.