INCH-POUND

A-A-54885
13 August 1993

COMMERCIAL ITEM DESCRIPTION

# CHART SET, VISION ACUITY TESTING, Clinical and Screening, Distant Vision

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-C-36228A, which is cancelled.

This Commercial Item Description covers a set of three alternate forms of charts for clinical and screening use in testing distant vision acuity.

Salient characteristics:

General. The vision test chart set shall consist of three alternate forms of test charts. Each chart form shall consist of two sections (charts), bound together with the spiral wire. Each section of each chart shall be made from three layers of cardboard, laminated by dry mounting process. Printing shall be by means of an offset process using the master negatives loaned by the Government to produce the printing plates. The outside corner of each chart form section shall be rounded. One 1/4 inch brass eyelet shall be on top centerline of each section.

Materials. Base weight of the cardboard shall be 300 pounds for 24 x 36 - 500 (plus or minus 5 percent). Caliper shall be 0.0203 inch. Bursting strength shall be 250 per square inch minimum. Tear strength shall be 650/880 minimum. Brightness shall be 85 minimum on 100 percent bleached Kraft stock.

The ink shall be a jet black offset ink.

The spiral binding wire shall be 0.0285 inch in diameter nominal, coiled to 7/16 inch outside diameter and approximately 6 coils per inch.

Dimensions. Each finished form (folded) shall measure approximately 22-1/2 inches wide by 23-1/2 inches long, excluding the binding spiral.

AMSC N/A FSC 6540

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Form Grouping. The loaned out Government negatives consists of six sheets. They are intended to make three alternate chart forms, as follows:

| Form No. | Sheet with Top Line Letters | mate with | Sheet with Top Line Letters |
|---|---|---|---|
| 1 | NS | | ZVDSO |
| 2 | CZ | | HZDCV |
| 3 | KD | | CSVZN |

Scoring instructions. The scoring instructions and applicable identifying form number are an integral part of the first half negative (lower left hand side) of each alternate chart form and shall be reproduced with test letters on the front of the chart section.

Workmanship. The charts shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). Each chart form shall be wrapped in suitable paper. The three wrapped forms comprising the unit shall be packaged in a suitable fiberboard carton or shall be wrapped and sealed in suitable flexible corrugated fiberboard. One set containing 3 wrapped forms, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

| National Stock Number | Item Identification |
|---|---|
| 6540-00-598-8075 | CHART SET, VISION ACUITY TESTING, Clincal and Screening, Distant Vision |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one chart set as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one chart set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commerical Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6540-1263
Location: c:harry\chart



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER<br>A-A-54885 | 2. DOCUMENT DATE (YYMMDD)<br>930813 |
|---|---|---|

3. DOCUMENT TITLE
CHART SET, VISION ACUITY TESTING, Clinical and Screening, Distant Vision

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| NAME | ORGANIZATION |
|---|---|
| ADDRESS (Include Zip Code) | TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON<br>(If applicable) |
| | DATE SUBMITTED (YYMMDD) |

7. PREPARING ACTIVITY

| a. NAME<br>Defense Personnel Support Center<br>ATTN: DPSC-MSE | b. TELEPHONE (Include Area Code)<br>(1) Commercial: (215) 737-2870<br>(2) AUTOVON: 444-2870 |
|---|---|
| c. ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.