INCH-POUND

A-A-59151
January 6,1998
SUPERSEDING
MIL-C-3363D
26 October 1992

COMMERCIAL ITEM DESCRIPTION

CHEST, MEDICAL EQUIPMENT MAINTENANCE AND REPAIR TOOL KIT

The General Services Administration has authorized the use of this commercial item description as a replacement for all federal agencies.

1. SCOPE.

1.1 Scope. This commercial item description (CID) covers one size of molded plastic tool chest with a removable tray.

1.2 Intended use. The chest is intended to be used to contain and protect tools and equipment required for maintenance and repair of medical service equipment.

2. SALIENT CHARACTERISTICS.

2.1 Design. The chest shall consist of a body, a lid, a removable tray, and shall be equipped with two handles and have locking capability.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be used in improving this document should be addressed to: U.S. Army Natick Research, Development and Engineering Center, Natick, MA 01760-5018 by using the Standardization Document Improvement Proposal(DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 5140

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

2.2 Construction. The chest shall have the lid fit down on the body and be watertight.

2.3 Dimensions. The chest shall have the following approximate dimensions: 27 inches in length, 17 inches in width and 14 inches in height - the total usable volume of approximately three cubic feet. The tray should have the following approximate dimensions: 15 inches in width, 3 1/2 inches in height and 26 inches in length.

2.4 Materials. The materials used shall be able to withstand temperatures of -25°F to 120°F and not exhibit any evidence of deterioration or deformation when exposed to a field environment.

2.5 Load capacity. The chest shall support a weight of 200 pounds when fully loaded and lifted by the handles. The empty chest shall weigh less than 40 pounds.

2.6. Color ("government unique" requirement) (see 6.3). The chest shall be olive green in color conforming to color numbers 24097, 34094, 34095, 34096 or 34097 of FED-STD-595. Chests of a specified lot shall not exhibit any variations in shade from the specific color selected for that product lot.

2.7 Handles. The handles should be integrally molded into the chest on each side and be capable of supporting the chest when fully loaded for transport.

2.8 Lock. A locking device shall be incorporated into the molded chest.

2.9 Drop test. The chest shall be capable of being dropped when fully loaded (200 pounds) on the bottom corners from a height of two feet onto a smooth hard surface(e.g. wood) without structural damage, cracking or denting.

2.10 Workmanship. The molded plastic chests shall be complete, clean and free of scratches, dents, breaks, sharp edges and corners and deformities. All components and assemblies shall be free of dirt and other extraneous materials. Burrs, slivers rough die, tool and grinding marks, dents and cracks shall be unacceptable. There shall be no defects affecting serviceability, durability, operation or appearance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

2.11 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the tolerances specified using conversion tables contained in the latest revision of IEEE/ASTM SI 10, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

3. REGULATORY REQUIREMENTS

3.1 Contractor-recovered materials. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR).

4. QUALITY ASSURANCE PROVISIONS

4.1 Product conformance. The products provided shall meet the salient characteristics of this commercial item description; shall conform to the producer's own drawings, specifications, standards and quality assurance practices and be the same product that has been sold in the commercial marketplace for over one year, or is the same product that has successfully been delivered to the Government on a previous contract or purchase order. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

5. PACKAGING

5.1 Packaging. Preservation, packing, labeling and marking requirements shall specified in the contract or purchase order (see 6.1).

6. NOTES

6.1 Ordering Data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents (if applicable).

a. Title, number and date of this document.
b. Packaging requirements (see 5.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

6.2 Sources of documents.

6.2.1 Sources for non-government association documents.

IEEE/ASTM STANDARD

SI 10 - Standard for Use of the International System of Units (SI): The Modern Metric System

(Application for copies of IEEE/ASTM standards should be addressed to the Institution of Electrical and Electronics Engineers, Inc., 345 East 47th Street, New York, NY 10017, or American Society for Testing and Materials, 100 Barr Harbor Drive, West Conshohocken, PA 19428).

6.3 Government unique requirements. Whenever a "government unique" requirement is included in the title of a paragraph under "Salient Characteristics", it means that the requirement is something that is not normally offered to the commercial marketplace by the manufacturer(s) of the subject commodity.

6.4 Sources of supply. Manufacturers whose products are known to meet the requirements of this document are listed below; however, competition is not limited to these companies.

Stanley Proto Industrial Tools
14117 Industrial Park Blvd., N.E.
Covington, GA 30209

Global Industrial Equipment
22 Harbor Park Drive
Department K
Port Washington NY 11050

Jensen Tools, Inc.
7815 S. 46th Street
Phoenix, AZ 85044-5399

Flambeau Commercial
Market Group
100 Grace Drive
P.O. Box 610
Roanoke Rapids, NC 27870

Rubbermaid, Inc.
Wooster, OH 44691-6000

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodian
Army - GL

Review Activity
Army - MD

CIVIL AGENCY COORDINATING ACTIVITY:

GSA - FSS

Preparing activity:

Army - GL

(Project 5140-0119)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2. The submitter of this form must complete blocks 4, 5, 6, and 7.

3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	**1. DOCUMENT NUMBER** A-A-59151	**2. DOCUMENT DATE (YYMMDD)** 980106

3. CHEST, MEDICAL EQUIPMENT MAINTENANCE AND REPAIR TOOL KIT

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME *(Last, First, Middle Initial)*	b. ORGANIZATION	
c. ADDRESS *(Include Zip Code)*	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON *(if applicable)*	**7. DATE SUBMITTED** (YYMMDD)

8. PREPARING ACTIVITY

a. NAME U.S. ARMY NATICK RD&E CENTER	b. TELEPHONE *Include Area Code)* (1) Commercial (508) 233-5175 (2) AUTOVON 256-5175
COMMANDER, U.S. ARMY NATICK RD&E CENTER ATTN: SSCNC-WEF NATICK, MA 01760-5018	**IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:** DEFENSE QUALITY AND STANDARDIZATION OFFICE 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22401-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:04Z
Check the source to verify that this is the current version before use.