COMMERCIAL ITEM DESCRIPTION

SLIDE, MICROSCOPE, (Frosted End, Beveled Edge)

This Commercial Item Description covers a 3x1 inch precleaned bevel edged frosted end microscope slide, for use in laboratories.

Salient characteristics:

Shall be precleaned glass microscope slide with a frosted end on one or both faces, and beveled working edges.

Shall be suitable for use in laboratory procedures involving making thin blood smears. Shall be suitable for use as both a slide that holds a blood sample for smearing and as a slide used for making that smear directly on the surface of another microscope slide.

Shall also be suitable for mounting histological and cytological preparations.

Shall have overall dimensions of 75 long by 25 wide millimeters with a tolerance of + 1.0 millimeters on both dimensions. The allowable thickness range shall be from 0.85 to 1.05 millimeters (mm).

Shall be fabricated from clear transparent glass that is free from imbedded foreign material, bubbles, seeds, blisters, striae and internal cloudiness. The surfaces of the slide shall be free from scratches in critical areas which would affect its use.

Shall be a precleaned slide, ready for use and free of detergent or residue. The faces of the microscope slide shall exhibit a uniform dispersion of liquid and absence of smears when dipped into a beaker of distilled or deonized water, quickly withdrawn, and immediately observed.

To facilitate blood smear work, each short end of the slide shall have doublesided beveled edges (i.e., two bevels per edge). The beveled working surfaces shall be smooth and uniform to allow for parallel movement on the surface (face) of another slide being smeared with a blood sample. The working ends (i.e., 1 inch ends) shall be beveled to approximately a 45 degree angle with the tip in the form of a knife edge, or a flat or rounded surface. As an alternate, the 1 inch ends may be partially beveled and need not be further ground.

The edges along the 3 inch dimension shall be free of sharpness to prevent finger cuts and facilitate handling by the user. The corners of the microscope slide shall be square, rounded or beveled.

The microscope slide shall have a single or double sided frosted end, 17 to 21 mm in length, which allows legible markings to be applied with a No. 2 lead pencil or a ball point pen. The marking sector(s) (frosting) shall be sufficiently coarse to distinguish by touch. The marking sector(s) shall be able to hold the applied markings through normal use including exposure to cytology fixatives (water and alcohol based) and submersion in alcohol and xylene.

Workmanship. The microscope slide shall be free from defects which detract from its appearance or impair its serviceability.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing 72 slides as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and Tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-01-172-6518

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL ENGINEERING COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6640-1294

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.