A-A-50981
19 July 1985

COMMERCIAL ITEM DESCRIPTION

CEMENT, ZINC PHOSPHATE, DENTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers zinc phosphate cement for use as a base under dental restorations, or as a luting agent for cast dental restorations.

Salient characteristics:

The cement shall be type I, fine particle size, cement in accordance with ANSI/ADA Specification No. 8 for Zinc Phosphate Cement. The color of the set cement shall be light gray shade, or light yellow shade as specified in the procurement document for the applicable National Stock Number and Item Identification.

Workmanship. The cement shall be clean and free from defects which detract from its appearance or impair its function.

Unit. One package containing one ounce of the powder component and sufficient volume of the liquid component, each packaged in a suitable container, constitutes the unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling and inspection procedures. Sampling and inspection procedures of ANSI/ADA Specification No. 8 shall apply for each lot of cement furnished.

Distribution Statement A: Approved for public release; distribution is unlimited.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6520-00-514-2350 CEMENT, ZINC PHOSPHATE, DENTAL, Light gray, 1 oz.
6520-00-514-2450 CEMENT, ZINC PHOSPHATE, DENTAL, Light yellow, 1 oz.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

AGENT:
DLA-DM

Project No. 6520-1843

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.