A-A-50984
24 July 1985

COMMERCIAL ITEM DESCRIPTION

BOTTLE DROPPER (Prescription)

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification DD-B-00585b, Type I Style I.

This commercial item description covers a bottle with a screw cap closure containing a pipet and bulb for use in dispensing small or drop doses of liquid medicaments as well as eye, ear and nose preparations.

Salient characteristics:

The bottle assembly shall consist of a bottle, a screw cap without a liner, a pipet and a bulb.

The bottle shall be fabricated of amber or green glass that conforms to the United States Pharmacopeia (USP) requirements for chemical resistance and light transmission.

The dropper bottle shall be oval shaped or oval on one side and flat on the other side to allow full display of labels. The bottle shall be narrow mouthed with well defined, continuous threads to receive the screw cap. The height exclusive of closure shall be:

1. 2-3/16 to 2-5/16 inches for 1/4 oz. size (5.56 to 5.88 cm for 7 g).
2. 2-17/32 to 2-23/32 inches for 1/2 oz. size (6.43 to 6.91 cm for 14 g).
3. 2-7/8 to 3-1/8 inches for 1 oz. size (7.30 to 7.94 cm for 28 g).
4. 3-7/16 to 3-15/16 inches for 2 oz. size (8.53 to 10.00 cm for 56 g).

The total capacity of the dropper bottle shall be greater than the nominal capacity to allow for displacement caused by the pipet.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

The dropper bottles shall be provided with the screw cap closures in place.

The screw caps shall be fabricated from a molded plastic material with sidewalls of adequate strength and weight for the purpose intended. The inner edges of the cap shall be threaded with a continuous thread formed as an integral part of the cap to mate and form a tight fit with the finish of the bottle.

The screw cap shall contain a centrally located hole for placement of a dropper assembly consisting of a pipet with attached bulb. The cap with the dropper assembly shall fit the bottle to form a liquid tight seal.

Pipets shall be fabricated from clear glass tubing, or from natural colored plastic material. Plastic pipets shall be nontoxic, in accordance with FDA food additive regulations, Food and Drug Administration, Code of Federal Regulations, Title 21, Food and Drugs, Parts 1 to 129. Plastic pipets shall show no decomposition or deformation when exposed to weak acids, in accordance with The American Society for Testing and Materials Designation D 543.

The pipet shall have a straight, tapered or blunt point tip and shall fit snugly into the rubber bulb.

The bulbs shall be made of rubber and shall have a sealing flange of uniform thickness and diameter to match the inside diameter of the screw cap.

Workmanship. Workmanship shall be first class throughout. The dropper bottles, screw caps, pipets and bulbs shall be free from defects which detract from their appearance or may impair their serviceability. All components shall be supplied clean, dry and free of dirt or foreign matter.

Unit. Each (EA) for 1/4 oz. bottle, or one box (BX) containing twelve 1/2- or 1 oz. bottles, or one box (BX) containing seventy-two 2 oz. bottles, as specified, constitutes one unit.

Reference standard. Where standard is referenced herein, the issue of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Contractor certification cont'd. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are representative of the items covered by this document:

6530-00-404-7500	1/2 oz (14 g)
6530-00-404-7700	1 oz (28 g)
6530-00-502-3500	1/4 oz (7 g)
6530-00-914-0239	2 oz (56 g)

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-50984

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1568

PGC 05475

Location: CID/S3-4-4
NCR disk, 7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50984	Bottle, Dropper (Prescription)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.