A-A-50987
2 August 1985

COMMERCIAL ITEM DESCRIPTION

BAND, COPPER, DENTAL

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification, RR-B-0096C (DSA-DM).

This commercial item description covers copper bands, suitable for use as a matrix in dental prosthetic procedures.

Salient characteristics:

Band shall be made of seamless, unannealed, electrolytic copper having a purity of at least 99.90 percent, with any silver counted as copper. Band shall be suitable for use as a matrix to confine dental impression materials in dental restorative procedures. Band shall measure 1/2 inch in height with a plus tolerance of 1/32 inch and a minus tolerance of 1/64 inch. Wall thickness shall be 0.0050 ± 0.0015 inch. Inside diameter shall be as indicated in Table I for the applicable size with a tolerance of ± 0.005 inch.

TABLE I. Inside Diameter of Bands.

Size	Inside Diameter (inch)
5	0.250
6	0.266
7	0.281
8	0.297
9	0.313
10	0.328
11	0.344
12	0.359
13	0.375
14	0.391
15	0.406
16	0.422
17	0.438
18	0.453
19	0.469
20	0.484

Band shall have a hardness of R(15T) 85 to 90 on the Rockwell superficial hardness scale. Band shall have a bright finish with no discoloration, smooth surfaces and clean cut edges free of burrs.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Workmanship. Band shall be bright, clean, smooth and free of dents and other defects which may detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing twelve bands of the size specified constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more test characteristics. The unit of product for inspection purposes shall be one band. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6520-00-501-5285	BAND, COPPER, DENTAL, SIZE 5, 12s
6520-00-501-5286	BAND, COPPER, DENTAL, SIZE 6, 12s
6520-00-501-5287	BAND, COPPER, DENTAL, SIZE 7, 12s
6520-00-059-9007	BAND, COPPER, DENTAL, SIZE 8, 12s
6520-00-059-9008	BAND, COPPER, DENTAL, SIZE 9, 12s
6520-00-059-9009	BAND, COPPER, DENTAL, SIZE 10, 12s
6520-00-059-9010	BAND, COPPER, DENTAL, SIZE 11, 12s
6520-00-059-9011	BAND, COPPER, DENTAL, SIZE 12, 12s
6520-00-721-6057	BAND, COPPER, DENTAL, SIZE 13, 12s
6520-00-059-9012	BAND, COPPER, DENTAL, SIZE 14, 12s
6520-00-985-7251	BAND, COPPER, DENTAL, SIZE 15, 12s
6520-00-720-9499	BAND, COPPER, DENTAL, SIZE 16, 12s
6520-00-059-9013	BAND, COPPER, DENTAL, SIZE 17, 12s
6520-00-059-9014	BAND, COPPER, DENTAL, SIZE 18, 12s
6520-00-059-9015	BAND, COPPER, DENTAL, SIZE 19, 12s
6520-00-059-9016	BAND, COPPER, DENTAL, SIZE 20, 12s

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians

Army - MU
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINTING ACTIVITY:

VA-OSS

Project No. 6520-1793

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.