A-A-50989
3 October 1985

COMMERCIAL ITEM DESCRIPTION

TUBE, COILED, BLOOD WARMING, (Sterile)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-T-36682A.

This commercial item description covers one type of disposable, plastic, coiled blood warming tube with a sterile fluid path suitable for warming blood from a blood bank from 35° to 98°F prior to transfusion.

Salient characteristics:

Shall be made of polyvinyl chloride plastic and rubber suitable for the intended purpose. None of the components coming in contact with the fluids being administered shall contain more than 5 parts per million of lead. The plastic shall meet the requirements of USP Class VI plastics. The unit shall be a plastic, coiled tube suitable for the transfusion of blood. There shall be a permanently affixed male Luer taper adapter at one end of the tube and a female Luer adapter at the other end. Each adapter shall be in accordance with ANSI/HIMA MD??C.1 and covered with a removable plastic protector. There shall be a Y-joint about 6 inches from the male Luer adapter. The free arm of the Y-joint shall have a rubber stopper reseatable to puncture by an 18 gage needle. A control clamp shall be present between the Y-joint and the coil of tubing. The unit shall be made to be placed in a water bath to warm blood from 35° to 98° F prior to transfusion. The flexible polyvinyl chloride tubing shall be 27-1/2 plus or minus 4 feet long, have an ID of 0.120 plus or minus .005 inch and a nominal OD of 0.170 inch. It shall have a fluid capacity of 58 plus or minus 9 ml. Loops of the coil shall adhere firmly to one another and the coil shall have an inside diameter of approximately four inches. The tubing shall be able to be pulled out of the coil without tearing the tubing. The female Luer adapter end protector may contain a cotton microbiological barrier which allows passage of gas.

The control clamp, if of the screw type, shall be made so that screw removal is difficult. The control clamp shall provide a constant setting and rapid shut-off from the open condition, shall be movable, and capable of one-handed operation. The control clamp when closed from the open condition ten times shall not cut, pierce, or damage the tubing.

The fluid path and the areas under the end protectors shall be sterile. Testing for sterility shall be accomplished by having an inoculum present with the product during sterilization and the testing in accordance with the USP method for viability of the microorganism in the inoculum. The microorganism shall be selected for its known resistance to the method of sterilization. The blood warming coil shall be non-pyrogenic when tested in accordance with the USP or an FDA and DPSC approved Limulus Amebocyte Lysate test method.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

The blood warming coil shall withstand without leakage 45 psig water pressure for 30 minutes or immersion in water at 10 psig air pressure for 10 seconds.

The Luer adapters and Y-joint shall be able to withstand a tensile stress of 4 pounds for 15 minutes without evidence of separation.

Unit. Each (EA). One tube, as specified, constitutes one unit.

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tube

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE The following National Stock Number (NSN) is covered by this document:

6515-00-926-2205

This item is intended to be used in conjunction with NSNs 6515-00-089-3470 and 6515-00-956-7795 - Blood Recipient Sets, Indirect Transfusion.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-50989

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

46510

Project No. 6515-2289

Location Chc S1-1-26

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.