A-A-50991
4 October 1985

COMMERCIAL ITEM DESCRIPTION

PAD, MAGNETIC, SURGICAL INSTRUMENT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers one type of flexible rubber instrument pad with imbedded magnets, suitable for holding steel alloy surgical instruments.

Salient characteristics.

Shall be a flexible pad with imbedded magnets designed to hold steel and steel alloy surgical instruments. The pad shall consist of 4 grommets, 54 rectangular magnets, and 2 flexible sheets.

Materials. The components of the pad shall be fabricated from the following materials: the flexible sheets shall be fabricated from natural rubber; the magnets shall be fabricated from any suitable material exhibiting magnetic properties; the grommets shall be fabricated from a suitable corrosion-resisting steel, properly hardened and passivated.

Construction. The magnets shall be arrayed in a square matrix with two rubber sheets molded over the magnets to secure them into a pad configuration. The pads shall possess permanently secured grommets at each corner.

Dimensions. The pad shall be 14 ± 1/32 inches long by 14 ± 1/32 inches wide by 1/4 ± 1/32 inch thick.

Weight. The total weight of the pad shall be not greater than 2 pounds.

Physical characteristics.

The pad shall show no evidence of deterioration when tested in accordance with the Thermal Stability Test cited below.

Thermal Stability Test. Sample pads shall be subjected to cyclic steam autoclaving at 270° ± 5°F, dry and saturated. A total of 10 cycles, 30 minutes ± 30 seconds per cycle, shall be used. After each cycle is completed, pads shall be examined for signs of swelling, tackiness, cracking, warpage, delamination, or staining.

Workmanship. The surgical instrument pad shall be free from any defects that detract from its appearance or impair its serviceability.

Unit. Each (EA). One pad, as specified, constitutes one unit.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror contractor shall comply, and be responsible for compliance by its subcontractors suppliers with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-163-3777	PAD, MAGNETIC, SURGICAL INSTRUMENT, FLEXIBLE, 14 by 14 inches (35.56 by 35.56 cm).

MILITARY INTERESTS:

Custodians

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

59746

Project No. 6515-3281

Location S1-8-4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.