A-A-50992
3 October 1985

COMMERCIAL ITEM DESCRIPTION

OTOSCOPE (BATTERY TYPE)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-O-36134.

This commercial item description covers a battery-type otoscope suitable for use in examination of, and surgery within, the ear.

Salient characteristics:

Shall be an otoscope assembly suitable for use in examination of and in operation upon the ear. Shall consist of the following components:

a) Otoscope head with magnifier, base connector and lamp.
b) Battery handle sized to hold two (2) D cells, without batteries.
c) Three (3) aural specula - 1 large, 1 medium and 1 small.
d) One (1) nasal speculum.
e) Carrying Case.
f) Two (2) replacement lamps.

Head to handle connection shall be a bayonet type connector and indexing lock mechanism to permit rapid attachment of the head and battery handle. The lamp shall have a minimum of 9 hours at 2.5 volts DC.

Top and bottom sections of the carrying case shall be inseparable and have a latch or other suitable means of closure.

The otoscope head shall contain lamp housing, magnifier, base connector and specula. The head shall contain a condensing lens if one is not integral with the bulb.

Specula shall be made of nylon or polypropylene with an interior treated or fabricated to minimize reflection. Small end shall have rounded edges.

Workmanship. The battery-type otoscope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One battery-type otoscope with case, as specified, constitutes one unit.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be a battery type otoscope.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-550-7200

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA - OSS

46002

Project No. 6515-3268

Location S1-1-40

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.