A-A-50993
4 October 1985

COMMERCIAL ITEM DESCRIPTION

MIRROR, LARYNGEAL

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-M-416.

This commercial item description covers a laryngeal mirror used for viewing the larynx and surrounding areas of the throat.

Salient characteristics:

Shall be a laryngeal mirror suitable for medical use in observing the larynx and surrounding areas of the throat. Shall have an approximate overall length of 8-1/2 inches, mounting a plain glass mirror on the distal end at an angle of approximately 35 degrees.

The laryngeal mirror shall be provided in the following sizes:

Size	Mirror Diameter (Inches approx.)
00	7/16
1	5/8
2	11/16
3	3/4
5	15/16

The mirror shall be circular, of uniform thickness; having front and back surfaces flat and even. The back of the mirror shall be thoroughly silvered and sealed to protect the coating. The mirror shall be securely mounted in the mirror holder in such a manner to protect the mirror from damage.

The laryngeal mirror shall show no evidence of corrosion or flaking, blistering, peeling or other damage to the silvered mirror surface or loosening of the mirror when boiled in distilled water continuously for 24 hours and dried for one hour at 230°F.

Shall be fabricated of a suitable metal plated with corrosion-resisting metal or fabricated of corrosion-resisting steel properly passivated. The mirror shall be high quality, polished glass, free from waviness and of such composition to withstand prolonged use and frequent heat and chemical sterilization, without dulling surface.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Workmanship. The laryngeal mirror shall be free from defects which detract from its appearance or impair its serviceability. The mirror and handle shall be symmetrically made, smoothly finished and free from burrs, dents, roughness, corrosion, dirt, grease, foreign matter and pitting. The mirror shall be free from haze and scratches and shall reflect a clear, undistorted image.

Unit. Each (EA). One mirror of the size specified, constitutes one unit.

Quality Assurance Provision

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances o procedures) and testing (determination by technical means of physical and chemical properties, of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one laryngeal mirror.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number(s) and applicable size are covered by this document.

National Stock No.	Size
6515-00-347-8400	00
6515-00-347-8500	1
6515-01-166-9537	2
6515-00-347-8600	3
6515-00-347-8700	5

Ordering data (Intermediate/exterior package quantities and labeeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-50993

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING AGENCY:

VA - OSS

05297

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Project No. 6515-3276

Location S1-3-16

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.