A-A-0050994A(DM)
25 July 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY-TYPE

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information. This document may be used in preference to A-A-50994 dated 4 October 1985.

Salient characteristics:

Except as specified below, shall be in accordance with American Society for Testing and Materials (ASTM) Designation: F 623-81, Foley Catheters, for requirements, test methods, and style and design.

The catheters shall be fabricated from soft, durable natural latex rubber. The entire lumen and external surface of the catheter shall be bonded with a layer of silicone rubber. The valve shall be fabricated from a rigid plastic compound.

The catheters shall be constructed with an integral self-retaining balloon. The capacity of the balloon, inside diameter and the breaking strength shall be as specified in Table I for each respective French size.

Table I. Dimensions and Physical Properties.

French Size (FR)	Capacity of Balloon (ml)	Inside Diameter Minimum (Inches)	Breaking Strength (Minimum pounds)
16	5	0.084	58
18	5	0.098	70
22	30	0.140	94
24	30	0.171	106

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Sterility. The catheters shall be sterile when tested in accordance with USP.

Workmanship. The item shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing ten or twelve catheters constitutes one unit. The catheters shall be individually packaged sterile in a peel-open type container capable of maintaining sterility of contents.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for tests. Sampling shall be conducted in accordance with MIL-STD-105, inspection level S-1, AQL 1.0. The unit of product for test purposes shall be one catheter.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:49Z
Check the source to verify that this is the currentversion before use.

Physical properties. Two straight specimens shall be cut from each test unit (catheter). One specimen shall be tested as cut. The other specimen shall be subjected to accelerated aging in accordance with Method 7221 of Federal Test Method Standard 601.

The specimens shall be tested for breaking strength and ultimate elongation in accordance with Methods 4111 and 4121 of Federal Test Methods 601 respectively, except paragraphs 5.2 and 5.3 of each Method shall be disregarded. The value of breaking strength shall be the value of the breaking force measured by Method 4111.

Unaged specimens shall have a minimum breaking strength as shown in Table I and an average ultimate elongation (arithmetic mean) of 750 percent. No unaged specimen shall have an ultimate elongation less than 680%. The breaking strength and ultimate elongation of the aged specimen shall not be less than 75 percent of the values obtained for the unaged specimen cut from the same catheter.

Any break which can be shown to have been caused by the test machine grips shall be classified as an invalid test result, and a complete entire catheter shall be re-tested.

Performance. Catheters shall be tested for performance in accordance with the performance tests of ASTM Standard F623. As an alternate test method for the Balloon Integrity and Balloon Volume Maintenance Tests, catheters may be overinflated with 50 cc of air, submerged in water, and the entire inflation system checked for leakage. Sampling for performance tests shall be S-1 (minimum) with an AQL of 1.0.

As an alternate test method for shaft size, the shaft may be measured by any acceptable method without cutting the shaft.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-104-8694	CATHETER URETHRAL, SILICONIZED RUBBER, DISPOSABLE ROUND TIP FOLEY,. With 5 ml Balloon, 16 Fr, 10s or 12s, Sterile
6515-00-104-8695	CATHETER URETHRAL, SILICONIZED RUBBER, DISPOSABLE ROUND TIP FOLEY,. With 5 ml Balloon, 18 Fr, 10s or 12s, Sterile
6515-00-104-8696	CATHETER URETHRAL, SILICONIZED RUBBER, DISPOSABLE ROUND TIP FOLEY,. With 5 ml Balloon, 22 Fr, 10s or 12s, Sterile
6515-00-104-9006	CATHETER URETHRAL, SILICONIZED RUBBER, DISPOSABLE ROUND TIP FOLEY,. With 5 ml Balloon, 24 Fr, 10s or 12s, Sterile

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Suspersession data. This interim commercial item description shall be used in preference to A-A-50994 dated 4 October 1985 and MIL-C-36891 dated 7 January 1974.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 05116

Project No. 6515-3821

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 — 606-036/44772

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-0050994A (DM)	CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY-TYPE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.