A-A-50995
4 October 1985

COMMERCIAL ITEM DESCRIPTION

CATHETER, URETHRAL, PLASTIC, DISPOSABLE ROUND TIP, ROBINSON-NELATON TYPE

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-C-36583A dated 18 August 1976.

This commercial item description covers Robinson-Nelaton type urethral catheter suitable for use in catheterization procedures.

Salient characteristics:

General. The catheter shall be a sterile, disposable, plastic, urethral catheter, Robinson-Nelaton design, French size 8, 10, 12, 14, 16 or 18, suitable for introduction into the human bladder through the urethra for the withdrawal of urine.

Material. The catheter shall be fabricated from soft, pliable, clear or red opaque, thermo-sensitive medical grade plastic which shall not cause trauma or irritate body tissue. The material used shall meet the requirements for Class IV plastics, as stated in the current United States Pharmacopeia (USP). The lead content shall not exceed five (5) parts per million (ppm).

Characteristics. The catheter shall have a nominal overall length of 16 inches with a perfectly round smooth tip, smooth circular shaft having a uniform diameter and integral funnel end for making easy connections. It shall be hollow behind the tip and have two (2) satin smooth eyes through the wall of the catheter tube that shall be staggered axially near the closed end and be diametrically opposed to each other. The eyes shall not overlap nor be separated by less than 0.105 inch (0.266 cm) nor more than 0.290 inch (0.737 cm). Commercial tolerances shall apply.

Physical properties.

Breaking strength. The breaking strength of the catheters shall be as specified in table I.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:49Z
Check the source to verify that this is the current version before use.

TABLE I. Breaking force

Size	8	10	12	14	16	18
Breaking strength (lb) minimum	5.0	12.0	18.0	24.0	30.0	36.0

Elongation. Elongation shall be 100 percent minimum.

Hardness. Durometer hardness shall be 77 ± 3 Shore "A".

Sterility. The catheters shall be sterile when tested in accordance with the USP.

Workmanship. The item shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with the requirements specified.

Unit. Package (PG). One package containing one hundred catheters constitutes one unit. The catheters shall be individually packaged sterile in a peel-open type container capable of maintaining sterility of contents.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheter.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:49Z
Check the source to verify that this is the currentversion before use.

NOTE: The following National Stock Numbers and their respective French sizes are covered by this document:

NSN	Fr. Size
6515-00-145-0011	8
6515-00-400-3399	10
6515-00-400-3401	12
6515-00-400-5465	14
6515-00-400-5466	16
6515-00-400-5467	18

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY

VA - OCS

05236

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Project No. 6515-3255

Location H1-6-19

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.