A-A-50997
4 September 1985

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS, Sterile

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers sterile, non-toxic, disposable, radiopaque, pyrogen-free, intravenous catheter placement units, suitable for extended intravenous therapy for a single venipuncture.

Salient characteristics:

Materials. The indwelling catheter shall be fabricated from clear or translucent, pliant plastic suitable in all respects for the purpose intended. The catheter shall be radiopaque throughout its length. The catheter, and all plastic components coming in contact with intravenous medication and patient's blood shall contain no more than 5 parts per million of lead. The finished catheter shall be sterile, pyrogen-free, and meet the requirements of the USP for class IV plastics. The needle cannula shall be fabricated from austenitic corrosion-resistant steel. The needle hub shall be fabricated from a suitable corrosion-resistant alloy or from a suitable plastic meeting the requirements of the USP for class II plastics. The catheter Luer hub and flow control plug shall be fabricated from suitable plastics.

Construction. The unit shall permit aseptic insertion by a surgeon without assistance and without the necessity of hand scrubbing or wearing of sterile gloves. The unit shall consist of an outer metal needle securely fastened to a needle hub. The needle cannula length shall be 2 ± 1/8 inches. Each unit shall be provided with components necessary to protect needle before use and the catheter from the needle after venipuncture. The needle hub shall possess a groove or other feature that shall provide for good gripping technique during catheter insertion.

The radiopaque catheter proper shall make a free, sliding fit within the needle. Proximal end of plastic catheter shall have a plastic Luer lock hub permanently attached to the catheter. A thin gauge plastic sleeve, sealed at its proximal end, shall enclose the catheter assembly up to the needle hub. Sleeve shall be securely fixed to the plastic needle collar at the distal end. Plastic needle collar shall have a stiffening flange and an opening to form a firm but not removable fit on the outside diameter of the needle hub. The unit shall be provided with a flow control plug. The plug shall appropriately seat into the catheter Luer hub. The catheter hub shall pass gauging tests with the devices listed in ANSI/HIMA Standard MD70.1.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Catheter hubs shall fit the plug gauge specified in ANSI/HIMA Standard MD70.1 within the "GO-NO GO" limits. Catheters of 19 gauge shall withstand an axial tensile force of 2.0 pounds and catheters of 16 gauge shall withstand an axial tensile force of 4.0 pounds.

Catheters shall not exhibit leakage exceeding one clinging drop (hydrostatic test) or one clinging air bubble (pneumatic test) when tested with appropriate fixtures and taper test gauges. The criteria shall be as follows:

Hydrostatic test - 45 pounds per square inch gauge (psig) pressure applied for 10 minutes.

Pneumatic test - 15 psig applied for 5 minutes.

Instructions. An instruction sheet for proper use of the catheter placement unit shall be provided with each complete unit.

Workmanship. The catheter placement unit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty catheter and needle units of one NSN only, individually packaged sterile, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one package of catheter and needle units.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers (NSNs) and their respective lengths, catheter gauges and needle gauges are covered by this document:

National Stock Number	Catheter Length	Catheter Gauge	Needle Gauge
6515-01-147-9526	8 inches	19	17
6515-01-147-9527	8 inches	16	14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-50997

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVILIAN AGENCY COORDINATING ACTIVITY:

VA - OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA - DM

06216

Project No. 6515-3280

Location S1-3-63

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.