COMMERCIAL ITEM DESCRIPTION

CYLINDER, GRADUATED, LABORATORY, POLYPROPYLENE

This commercial item description shall be used in preference to Military Specification MIL-C-36229.

This commercial item description covers requirements for polypropylene graduated cylinders, suitable for laboratory purposes.

Salient characteristics:

Shall be cylindrical, graduated cylinders with supporting base, having the following capacities; 10 millimeter (ml), 25 ml, 50 ml, 100 ml, 250 ml, 500 ml, 1000 ml and 2000 ml.

Shall be fabricated of a good quality polypropylene material, and provided with a pour-out spout.

The base shall be octagonal and constructed in such a manner that the empty cylinder is prevented from tipping over when placed at an angle of 15^{0} from the perpendicular.

Shall be permanently and legibly calibrated in accordance with NBS Circular 602, Testing of Glass Volumetric Apparatus.

Main graduations shall be suitably inscribed in Arabic numerals. The graduation ranges and intervals shall be as shown in table I.

TABLE I. Graduation intervals.

Size	To be graduated between (inclusive)	Graduations Least	Intermediate	Main	Minimum width of intervals (millimeters)
10 ml	1 to 10	--	0.5	1	1.0
25 ml	5 to 25	0.5	1	5	0.9
50 ml	10 to 50	--	1	10	1.8
100 ml	10 to 100	1	5	10	1.2
250 ml	50 to 250	--	10	50	1.5
500 ml	100 to 500	5	25	100	2.4
1000 ml	100 to 1000	--	10	100	2.7
2000 ml	200 to 2000	20	100	200	3.5

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Graduation marking shall be raised, permanent, legible and a maximum of 0.6 millimeter wide.

Shall be calibrated to deliver the indicated volume @ 25°C (77°F).

Instructions shall be included with a warning that autoclaving might alter values of calibrations.

Workmanship. The cylinders shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One cylinder, of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection and testing. Inspection and testing shall be conducted to determine compliance with specification requirements.

Sampling for inspection and testing. Sampling for inspection and testing shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are representative of the items covered by this document:

6640-00-889-7087, 6640-00-889-7088, 6640-00-889-7089, 6640-00-889-7090
6640-00-889-7091, 6640-00-889-7092, 6640-00-889-7093, 6640-00-982-7492

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

Civilian Agency Coordinating Activity:

VA - OSS

Project No. 6640-1118

Loc. S3 7-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50999	CYLINDER, GRADUATED, LABORATORY, POLYPROPYLENE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording.

b. Recommended Wording:

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.