A-A-51000
8 August 1985

COMMERCIAL ITEM DESCRIPTION

RACK, TEST TUBE, LABORATORY

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-R-36922 (DSA-DM).

This commercial item description covers a twenty tube Wasserman metal test tube rack suitable for use with water baths.

Salient characteristics:

Shall be a twenty tube Wasserman test tube rack with 5/8 inch-0 + 1/64 diameter holes for use with water baths.

Shall be fabricated from, as a minimum, 0.025 inch thick No. 24 United States Standard (USS) gauge corrosion-resisting steel. Commercial tolerances on the thickness shall apply.

Shall be free from rocking and wobbling when resting on a flat surface.

Shall consist of two perforated flanged shelves, an unperforated flanged base, and two flanged supporting ends with cutouts for lifting and drainage. The flanged ends of the base and shelves shall be fitted inside the flanged supporting ends, and affixed by welding (e.g. spot welding).

The perforated shelves shall have two rows of ten holes that are uniformly spaced, and numbered in pairs from 1 to 10 lengthwise. The center lines of corresponding holes in the two shelves shall be in proper register to assure that the inserted tubes rest vertically in the rack.

The overall dimensions shall not exceed 10-1/8 long by 2-1/8 wide by 4-1/2 high (inches). The top shelf surface shall be at least 2-1/4 inches above the bottom level of the end supports, and the middle shelf surface shall be at least 1.0 inch below the top shelf surface.

Workmanship. The test tube rack shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test tube rack, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-00-442-6450

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract/and or order.)

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

A CIVIL AGENCY COORDINATING AGENCY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

55429

Project No. 6640-1199

Location: S3-4-5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.