A-A-51001
5 September 1985

COMMERCIAL ITEM DESCRIPTION

CATHETER, CHOLANGIOGRAM,
Plastic, Disposable, Sterile

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a catheter designed for intra-operative cholangiography.

Salient characteristics.

The catheter shall be sterile, disposable, radiopaque, 8 French, 40 cm length, specifically designed for intra-operative cholangiography. The catheter shall be fabricated from a flexible plastic, clear and transparent, for the visualization of air bubbles. The radiopaque ring shall be 1.5 cm from the tip opening for placement determination of the catheter end near the common duct.

The tube shall have a radiopaque stripe for its entire length except a portion of the distal end. The tube shall have a beveled distal tip simplifying catheter insertion into the cystic duct. The distal portion shall be cone-shaped with the flared proximal end of the cone 0.5 cm from the catheter tip. The distal portion shall be specifically designed to be held easily by an external suture and to prevent accidental withdrawal.

The cone flange in the cystic duct shall be designed to minimize leakage of contrast media. The catheter shall contain two parts equipped with luer-lock fittings to allow dye and irrigation syringes to be attached simultaneously.

The French size shall be clearly marked on the catheter.

The catheter shall meet any FDA requirements for safety and toxicity applicable for invasive and implantable devices in effect at the time of solicitation.

Workmanship. The catheter shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirement specified

Unit Each (EA) One catheter, package unrolled in a reel-open type container capable of maintaining sterility of content while the container is damaged or opened, constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection as used herein is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties of the item

Tests. Tests shall be conducted to determine compliance with specification requirements Where feasible the same sample shall be used for the determination of two or more test characteristics

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD- or as specified herein The [illegible] shall be stated

Sterility test. The [illegible] catheter shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, or paper spore strip (carrier Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution The minimum number of test units regardless of lot size, shall be 10

Alternate method. Suppliers conducting sterility tests shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the contracting officer, and approval obtained before employment of the alternate method on government contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Radiopacity. The cholongiography catheter shall be placed on an unexposed X-ray plate and covered with an aluminum top wedge ranging in thickness from 1/8 inch to 3/4 inch in steps of 1/8 inch. Using an appropriate exposure so that the steps of at least 1/4, 3/8, 5/8 inch are clearly differentiated from each other, the insert appearance shall be noted

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes and laws, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered materials in accordance with Public Law 94-580 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal material handling conditions, handling, etc involving shipment from the supply source to the receiving activity, plus redshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior packaging quantities and labeling and marking shall be as specified in the contract and/or order

NOTE The following National Stock Number (NSN) is covered by this document

[illegible]

Military Interests

Custodians:
Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activity.
VA-OSS

Preparing Activity
DoD-MB

Agent:
DLA-DM

PGC 60974

Project No. 6515-3326

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.