A-A-51003
30 August 1985

COMMERCIAL ITEM DESCRIPTION

BRUSH, SCRATCH, DENTAL

The General Services Administration has authorized this commercial item description in preference to Interim Federal Specification H-B-00248D (DSA-DM).

This commercial item description covers a flexible steel wire scratch brush suitable for cleaning dental burs, files, and instruments.

Salient characteristics:

Material. The sleeve shall be made of plastic tubing. The wire shall be fabricated from drawn, flexible, corrosion resistant steel, or AISI 1022 or 1008 carbon steel rendered rust resistant by a suitable galvanized coating.

Design and dimensions. The brush shall be a bundle of wires enclosed in the sleeve. The wires shall be completely bent in the middle so that the two halves of the wire are parallel. The bent wires shall be squeezed together to form a bundle which shall be inserted into the sleeve. All bent areas shall be at one end of the bundle. The loose ends shall be uniformly trimmed and there shall be no stray or improperly fitted wires. The sleeve shall serve as a hand grip. The edge of the proximal end shall be clean cut and the edge of the distal end shall be free of sharp projections. The sleeve shall be trimmable with a dental laboratory knife so that the metal surface may be exposed as the loose ends of the wire wear down. The scratch brush shall be of the dimensions shown in Figure 1.

Workmanship. The sleeve shall be free of nicks, cuts and tears. The scratch brush shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One brush, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.




FIGURE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one brush. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6520-00-503-1000 BRUSH, SCRATCH, DENTAL

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1852

☆U.S. GOVERNMENT PRINTING OFFICE: 1985 — 505-038/23957

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.