A-A-51004
30 August 1985

COMMERCIAL ITEM DESCRIPTION

GAGE, ENDODONTIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a measuring gage.

Salient characteristics:

The measuring gage shall be suitable for establishing and presetting the operating limit of endodontic instruments. The gage shall be of two-piece construction: a fixed piece and a sliding piece with a thumb screw to provide a positive locking mechanism along the fixed piece. The two pieces and screw shall be fabricated of austenitic stainless steel. The gage shall have two scales with a range of 0-35 mm, legibly and permanently numbered every 5 mm, with 1 mm markings, 1/16 inch high, nominal. The scales shall be placed on the same edge of the fixed piece and shall start at zero, beginning at each end. Each piece shall have a perpendicular end tab. The tab on fixed piece shall have a "V" notch for accurately positioning an instrument to preset its operating limit. Two holes shall be tapped on the centerline of fixed piece to accommodate round head, slotted, machine screws with appropriate size washers. One hole shall be tapped on centerline of slide for thumb screw.

The overall length shall be 4 ± 1/16 inches and 23/32 ± 1/32 inch wide, with a hole tapped 1-15/32 ± 1/32 inch from free end and a second hole 3/4 inch from the first hole. The tapped hole on slide shall be 27/32 ± 1/32 inch from free end.

The three tapped holes shall be #3-48 UNC-2B. The "V" notch shall be 3/8 ± 1/32 inch deep.

The gage, except two machine screws and washers, shall comply with the boil test of the latest issue in effect of ANSI/ADA Specification No. 29.

Workmanship. Gage shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One gage, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or

FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one gage. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6520-01-061-6445 GAGE, ENDODONTIC

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 05

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DH

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

49039

Project No. 6520-1849

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.