A-A-51006A
25 November 1985
Superseding
A-A-51006
4 September 1985

COMMERCIAL ITEM DESCRIPTION

CLAY, MODELING (Dental)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers modeling clay suitable for use in dental laboratory procedures.

Salient characteristics:

The modeling clay shall be suitable for use in dental laboratory procedures such as model block out and sealing duplicating flasks. The modeling clay shall be suitable for use in forming the land area on an impression prior to the boxing procedure.

Material. The modeling clay shall consist essentially of kaolin, calcium hydroxide and calcium carbonate incorporated in an oil-base consisting of microcrystalline wax, paraffin, petrolatum and color additives. Color of clay shall be optional. The clay shall be of a homogenous consistency free of lumps, grit or other impurities which may impair its use. The clay shall contain no ingredients harmful to the hands. The clay shall be free from unpleasant or irritating odors.

Working properties. The clay shall be sufficiently plastic to be easily hand molded at room temperature so as to produce a dry smooth surface capable of retaining its shape for a minimum of 24 hours without toughening, hardening, drying out, crusting or cracking. The clay shall be capable of an exposure of ten minutes to temperatures up to 200° Fahrenheit without change in shape or formation of gas.

Workmanship. The modeling clay shall be clean and free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One two-pound package of modeling clay, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one 2-pound package of modeling clay. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

6520-01-201-7029 CLAY, MODELING

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

58408

Project No. 6520-1884

Location S2-10-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.