A-A-51007
13 September 1985

COMMERCIAL ITEM DESCRIPTION

FLASHLIGHT, SURGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, disposable, self-contained, flexible, hand-held light suitable for use in minor surgical procedures.

Salient characteristics.

Shall be a sterile, disposable, self-contained, flexible, battery powered hand-held light. The flashlight shall be powered by two AA size alkaline batteries contained in a suitable rigid plastic housing. Shall have a flexible shaft 5 to 6 inches long fitted with a 270 to 300 milliamp, 2.3 to 3 volt pre-focused lamp at the distal end. The battery housing shall be provided with a suitable on-off switch designed to preclude the flashlight from accidentally being activated. The flashlight shall be provided with a cover to protect the flexible shaft and lamp from damage during handling, shipment and storage. The cover shall be easily removed, but shall fit snugly enough onto the battery housing to support its own weight in the vertical position. The flexible shaft shall be securely fastened to the battery housing and shall be completely covered with an insulating material. The lamp and its connec- tion shall be suitably attached and shall be reinforced to prevent breakage of con- nection or lamp during normal usage. The expiration dating period shall be not less than 24 months. Shall be sterile when tested in accordance with the requirements of the United States Pharmacopeia (U.S.P.).

Expiration dating period:

Not less than six months of the expiration dating period shall remain at the time of delivery of the flashlights to the Government.

Tests

Tests to show the condition of fresh batteries shall be started within 30 days after shipment by the battery manufacturer. Manufacturer shall select the applicable test from American National Standard Specifications for Dry Cells and Batteries (ANSI C18.1-1979). The test used shall be the one that most nearly duplicates the power output of the battery in actual use.

Workmanship. Flashlights shall be free from defects which detract from their appearance or impair their serviceability.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing ten flashlights, each flashlight individually packaged sterile in a sealed, peel-open type container constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6515-00-149-1402

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PSC 46049

Project No. 6515-2597

U.S. GOVERNMENT PRINTING OFFICE: 1985 - 505-038/24104

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.