A-A-51008
10 October 1985

# COMMERCIAL ITEM DESCRIPTION

## CAUTERY, BATTERY OPERATED

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers two types of compact, fully-sealed, sterile disposable cauteries for control of small bleeding sources in minor surgery.

Salient characteristic.

Shall be a fully-sealed, sterile, disposable, battery-powered cautery. The tip shall have an operating temperature of 2100°F (1148°C) ± 10 percent. Shall have maximum dimensions of 7 inches overall length and 3/4 inch diameter. Weight shall be 2-1/2 ounces ± 1/2 ounce. Case shall be fabricated of a suitable plastic, shall house two AA size alkaline batteries (ANSI designation L-40). Shall be furnished with an actuator button capable of immediate start and stop control and a tip overcover; actuator button shall be inoperable when tip overcover is in place. Combined length of tip and shaft shall not exceed 4-1/2 inches. The Cautery Fine Tip shall be approximately 5/8 inch (1.58 cm). The Cautery Standard U-Tip shall be approximately 1 inch. Not more than 6 months shall have elapsed from the date of manufacture of the cautery to the time of delivery to the Government. Batteries shall not be older than 6 months at time of delivery

Cautery shall be furnished in two types

Type I - U (loop) tip
Type II - Fine tip

Tests

Tests to show the condition of fresh batteries shall be started within 30 days after shipment by the battery manufacturer. Manufacturer shall select the applicable test from American National Standard Specifications for Dry Cells and Batteries (ANSI C18.1-1979). The test used shall be the one that most nearly duplicates the power output of the battery in actual use

Workmanship. The cauteries shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing ten cauteries, individually packaged sterile, of one type only, as specified, constitutes one unit.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

The following National Stock Numbers are covered by this document

6515-01-021-1361 (Type I)
6515-01-021-3688 (Type II)

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY.

VA-OSS

PGC 05831

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6515-2570

Location CID H1-2-23

GU S GOVERNMENT PRINTING OFFICE 1985 — 605 036/24793



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.