A-A-51010A
25 October 1989
SUPERSEDING
A-A-51010
10 October 1985

COMMERCIAL ITEM DESCRIPTION

VIEWER, DENTAL RADIOGRAPHIC FILM

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-V-350 which is cancelled.

This Commercial Item Description covers a viewer suitable for viewing dental radiographic films.

Salient characteristics:

The dental radiographic film viewer shall be suitable for viewing dental radiographic films, with or without mounts, as well as general purpose x-ray and photoroentgenographic films up to and including 5-1/4 by 13-1/4 inches. The viewer shall have the following overall dimensions:

Length: 16 inches plus or minus 1 inch

Height: 6 inches plus or minus 1 inch

Depth: 3 inches plus or minus 1 inch

The viewer shall be of fabricated cold-rolled steel or corrosion resistant steel. All metal parts, with the exception of stainless steel and plated steel, shall be finished with a suitable enamel. The viewing area shall be a plastic diffusing panel not less than 5-1/4 by 13-1/4 inches, capable of accepting a full mouth series mount of at least 18 windows. The diffusing panel shall have a light transmission factor of 55 plus or minus 5 percent.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

The viewer shall be suitable for wall, shelf, or desk mounting. The viewer shall have stainless steel legs for desk or shelf mounting. The legs shall be removable or recessed for wall mounting.

The viewer light source shall be a daylight fluorescent lamp, rated at approximately 14 watts, for use with nominal 115 volt, 60 Hz, AC power. The viewer shall include an "on-off" switch, a three conductor line cord with ground lead, approximately 6 feet in length and a three prong, hospital grade, grounded male connector conforming with ANSI Standard C73.11.

All electrical wiring shall conform to National Fire Protection Association (NFPA) Publication No. 70, National Electric Code. The ballast, lamp, switch, electrical cord, and plug used in the viewer shall conform to the applicable requirements of Underwriter's Laboratories, Inc. The illuminator shall show no signs of dielectric breakdown when tested for dielectric withstand as specified in UL 153.

The component parts of the viewer shall be securely fitted and properly assembled. The parts shall be free of cracks, tears, ruptures, or ripples. The viewer edges shall be free of burrs and sharpness.

The viewer shall be free of unplated or unpainted areas. The plating or enamel shall be free of blisters.

Workmanship. The dental radiographic film viewer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One viewer, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for dimensional examination shall be conducted in accordance with MIL STD-105 with an AQL of 2.5 (percent defective) and an inspection level of S-3. Failure of the dielectric withstand test shall constitute rejection of the tire lot. The unit of product for test purposes shall be one dental radiographic film viewer.

Tests

Dielectric withstand. The viewer shall be tested for dielectric withstand as specified in UL 153. Any sign of dielectric breakdown shall constitute a failure.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-455-9947

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1254

Location:
CID/VIEWER/DENTAL/RADIOGRAPHIC/
FILM/H13-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-51010A	VIEWER, DENTAL RADIOGRAPHIC FILM

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*:
b ADDRESS *(Street, City, State, ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.