A-A-51012A
25 October 1989
SUPERSEDING
A-A-51012
10 December 1985

COMMERCIAL ITEM DESCRIPTION

PAD, HEAT TREATMENT

(Moist Heat)

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-37058 which is cancelled.

This Commercial Item Description covers a moist heat treatment pad suitable for use in the treatment of muscular spasms and pain due to neuromuscular conditions.

Salient characteristics:

The heat treatment pad shall be a moist heat therapy pad suitable for use in the treatment of injuries, arthritis, and other neuromuscular conditions. The pad shall be made in the following styles:

Type I - Rectangular style

Type II - Neck contour style

The type I pad shall be 9-3/4 inches plus or minus 3/4 inch wide by 11-1/2 inches plus or minus 3/4 inch long. The pad shall be stitched quilt-like into 12-14 rectangular sections (pockets) to keep the hydrous sodium aluminum silicate filler properly distributed within the pad, each pocket containing silicate granules. The pad shall have a loop for handling attached at each of its four corners.

The type II pad shall be 24 inches plus or minus 1 inch long with one edge straight, the other a double reverse curve, so that the width at the ends is 3-1/2 inches plus or minus 1/2 inch and the width at the center is 6-1/2 inches plus or minus 1/2 inch. The pad shall be stitched quilt-like into 12-16 sections (pockets) to keep the hydrous sodium aluminum silicate filler properly distributed within the pad, each

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

pocket containing silicate granules. The pad shall have a loop for handling attached at each end of the straight side.

Cloth for the pad shall be a suitable cotton cloth. The weave and weight of the cloth shall be tight enough to retain the silicate when hydrated, but still porous enough to permit easy escape of steam from the heated hydrated filler. All sewing shall be machine stitching with 8 to 14 stitches per inch. The pad shall be free from raw or raveled edges. The pad stitching shall be backstitched at the ends, or at thread breaks, to prevent raveling. All seams shall be straight and evenly spaced.

The silicate granules shall absorb a minimum of 4 times their weight of water, and at full saturation shall occupy a volume 10 to 15 times their dry bulk. A failure to achieve either of the above when tested as specified shall constitute a defect.

The pad shall have a printed or woven label securely stitched to it. The label shall bear the name or registered trademark of the manufacturer. The pad marking shall be with waterproof ink, waterproof paint, or woven lettering which will not be affected by continuous water immersion at 160°F which the pads normally undergo.

The pad shall be strong enough to withstand repeated manipulation and flexing while in the swollen condition and shall show no evidence of broken threads, torn cloth, or loss of silicate granules during use.

The pad shall be clean and free of dirt, oil, grease, or other contamination.

Workmanship. The heat treatment pad shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One pad of the type specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-3. The unit of product for test purposes shall be one heat treatment pad.

Tests

Silica gel absorption test. Steadily pour a measured, weighed amount of silica gel into a beaker containing a measured amount of cold water. After the gel swells, forming a smooth, creamy paste, the solution shall be inspected. Failure for the silica gel to absorb at least 4 times its weight of water, and at full saturation occupy a volume 10 to 15 times its dry bulk, shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

6530-00-711-3010 (Type I)
6530-00-865-2288 (Type II)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
HHS
FDA-MPQAS

Project No. 6530-2091

Location:
CID/PAD/HEAT/TREATMENT/H13-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-51012A	PAD, HEAT TREATMENT, (Moist Heat)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*.

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 82 MAR PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.