A-A-51015
20 September 1985

COMMERCIAL ITEM DESCRIPTION

CUP, SPECIMEN

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-C-36542.

This commercial item description covers a disposable plastic specimen cup with lid, suitable for use in military medical treatment facilities.

Salient characteristics:

Shall be a sterile, disposable, translucent, polystyrene or polypropylene container of 4-1/2 ounce (133 ml) capacity, suitable for the collection of laboratory specimens. Shall be supplied with a sterile screw cap.

The container shall be circular in section, approximately 2-1/2 inches (6.4 cm) in diameter and approximately 2-1/2 inches (6.4 cm) in height. Shall be graduated in 1/4 ounce (7.4 ml) steps over a minimum range of 3/4 ounces (22.2 ml) to 4 ounces (118 ml), or in 10 ml steps over a minimum range of 20 ml to 110 ml. Graduation numbers or lines shall be clearly defined and easily read.

The screw cap shall fit the container securely, be easily assembled and form a tight seal to prevent spillage of contents.

The full container shall be stable when placed on a flat level surface.

Container and cap shall be free from objectionable odor, sharp edges, deformations, cracks, burrs, holes, or thin areas.

One container and one cap together shall be sealed in an individual clear plastic container or envelope of a peel-open type. This plastic container or envelope shall maintain the sterility of the contents until manually opened. It shall also permit aseptic handling of the cup and delivery of the cup, uncontaminated, onto a sterile field. A tear-open type package shall not be acceptable. The label of the immediate container shall bear the word "sterile".

The plastic container or envelope shall be strong enough to avoid holes, perforations, or tears during shipment or handling.

Labeling shall be that usually furnished by the supplier as commercial practice. Addition to these labels shall be limited to application of the following to the unit of issue (strip labeling is acceptable):

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

a) National Stock Number (NSN).

b) Date of packaging (month and year, i.e., 4/75). Final sterilization shall be on or after packaging date.

When the unit package is the shipping container, additional labeling may be applied.

Workmanship. The plastic specimen cup shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred specimen cups, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

The following National Stock Number is covered by this document:

6530-00-837-7472

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 59809

Project No. 6530-1786

location H1-12-14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.