A-A-51016
10 October 1985

# COMMERCIAL ITEM DESCRIPTION

## TRAY, CATHETER

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification RR-T-00626.

This commercial item description covers stainless steel catheter trays, with covers, suitable for holding catheters used in military medical treatment facilities.

Salient characteristics:

Shall be a catheter tray with cover and handle designed to hold catheters in hospital operating rooms and in medical facilities. Shall be made in the following nominal sizes (inside dimensions):

Small - 8 inches (20.3 cm) long by 2-1/2 inches (6.4 cm) wide by 1-1/2 inches (3.8 cm) deep.

Medium - 16-1/2 inches (41.9 cm) long by 3-1/2 inches (8.9 cm) wide by 2-1/4 inches (5.7 cm) deep.

The tray and cover shall be formed seamlessly, each from one piece of 0.031 inch thick (minimum) corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series, which shall not be reduced below 0.025 inch thickness at any point. (The cover of the small tray shall be formed from 0.025 inch thick (minimum) steel which shall not be reduced below 0.021 inch thickness at any point). The cover handle shall be formed of 0.062 thick (minimum) corrosion-resisting steel AISI 300 series.

The bottom of the tray shall be so designed that maximum reinforcement and rigidity is obtained. The tray shall rest on a level surface without rock or wobble.

Cover shall be designed to fit tray squarely, without rock or tendency to rattle. Cover may be flat or domed.

Cover handle shall be square or rounded, provided finger clearance is adequate.

Spotwelding of the handle to the cover shall be firmly secure, smooth, and free of cracks, porosity, and other defects.

Tray and cover surfaces shall have a uniform #4 commercial finish, or better, and be readily cleanable.

Tray and cover shall be free of cracks, sharp edges, burrs, pits, dents, ripples, and deformities.

FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Workmanship. The catheter tray shall be free from defects which detract from its appearance of impair its serviceability.

Unit. Each (EA). One tray with cover, of the size specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Numbers are covered by this document:

6530-00-793-8210 (small)
6530-00-793-8220 (medium)

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY.

VA-OSS

PGC 05149

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6530-1787

Location H1-13-2



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.