A-A-51018
10 October 1985

COMMERCIAL ITEM DESCRIPTION

BOTTLE, URINE SPECIMEN AND CAP, SNAP-ON

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a plastic, disposable urine specimen bottle with plastic snap-on cap suitable for use in military medical treatment facilities.

Salient characteristics:

Shall be an unpigmented, disposable polypropylene bottle of at least 6 ounce capacity with snap-on cap suitable for use in collection, transportation and analysis of urine specimens within military medical facilities.

BOTTLE

The bottle shall be approximately 4 inches in height, with a base approximately 2-1/4 inches square, or approximately 2-1/2 inches in diameter. Walls shall taper to form a round neck approximately 3/4 inch in height and 1-5/8 inch outside diameter. Neck shall have an integral non-drip pouring spout and an inner well for test strip dip testing of small volume urine specimens. Dip well shall be located opposite (180^{0}) from the pouring spout.

The bottle shall be distinctively graduated on a dual scale in 1/2 ounce steps from 1 to 6 ounces and in 15 milliliter (ml) steps from 30 ml to 180 ml. Scale shall be numbered at 1, 2, 3, 4, 5, and 6 ounce levels and at 60, 120, and 180 ml levels. Graduations shall be distinct and legible and labeled as mls or ounces.

The bottle shall be stable when placed on a flat level surface.

Rim shall be free of all deformities and imperfections.

CAP

The cap shall be a round, unpigmented plastic cap approximately 1-3/4 inches in diameter and 1/4 inch high.

The cap shall have a lift tab appropriately sized to provide for easy removal of the cap from the bottle.

Cap design shall include an annular bead on the inside to allow a snap-on fit on the urine specimen bottle. Cap shall protect contents from dust and other contaminants.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Bottle and cap shall not contain embedded dirt, unplasticized resin or other foreign objects.

Bottle and cap shall be free of objectional odor, sharp edges, cracks, holes, or thin areas.

Bottle and cap assembly shall be leakproof.

Workmanship. The urine specimen bottle and cap shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 500 bottles or one package containing 1000 caps, as specified, constitutes one unit of issue.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:49Z
Check the source to verify that this is the current version before use.

The following National Stock Numbers are covered by this document:

6530-00-080-0040 (urine specimen bottle)
6530-00-080-0049 (cap)

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

Va-OSS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PGC 05152

Project No. 6530-1789

Location H1-12-15

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.