A-A-51022
10 October 1985

COMMERCIAL ITEM DESCRIPTION

TRAY, MEDICINE

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification RR-T-00640.

This specification covers a stainless steel, rectangular medicine tray suitable for use in military medical treatment facilities.

Salient characteristics:

Shall be a rectangular tray, without a cover, and suitable for carrying catheter trays, thermometer technique equipment and medicines to a patient's bedside.

The tray shall be formed seamlessly from one piece of .031 inch thick corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series, which shall not be reduced below 0.25 inch thick at any point.

The tray shall have approximate dimensions of 13-1/2 inches long, 9-5/8 inches wide, and 5/8 inch deep. Upper lip shall be approximately 1/8 inch. Corner radius (at top of tray) shall be approximately 1-1/4 inches maximum.

The tray shall rest on all four bottom corners on a level surface without rocking or wobbling. The bottom may concave inward so that the highest point shall be approximately more than 1/16 inch above the surface upon which it rests.

Tray surface shall have a unifrom #4 commercial finish or better and be readily cleanable.

The tray shall be free of sharp edges, burrs, pits, dents, cracks, ripples, and deformities.

Workmanship. The medicine tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). Each tray, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6530-00-794-0600

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC 50441

Project No. 6530-1782

Location H1-11-5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.