METRIC

A-A-30099D

April 20, 1993

SUPERSEDING
A-A-30099C
November 27, 1991

COMMERCIAL ITEM DESCRIPTION

CATHETERIZATION KIT, URETHRAL

This commercial item description describes catheter trays, with and without catheter. It is a compact, disposable, sterile unit which includes essential supplies for a simple catheterization procedure.

Salient characteristics The urethral catheterization kit shall provide sterile disposable components arranged in proper sequence for catheterization technique The catheterization kit shall be provided in either Type I, II, III or IV kits as described herein. The kit shall consist of as a minimum, the following components

Component	Type I	Type II
Plastic tray 1 ea graduated in 100 ml increments	minimum 1000 ml capacity with 3 compartment lift out tray	700 ml capacity minimum
Absorbent cotton/ rayon balls	Minimum 5 ea, rayon 3 cm dia , min. minimum weight, .91 grams each. Cotton, 3.81 cm minimum, weight .3 grams	5 ea minimum approx. 3.81 cm diameter
Underpad, 1 ea	40 x 40 cm minimum	48 x 41 cm minimum
Fenestrated drape, 1 ea	33 x 34 cm minimum with 9.5 cm square approx. fenestration or 13.4 cm fenestration point to point diagonally	Approximately 46 x 33 cm, with 9.5 cm square approx. fenestration

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Component	Type I	Type II
Lubricant, 1 ea	Minimum 2.5 grams bacteriostatic or terminally sterilized	5 grams water soluble with or without topical anesthetic
Specimen Container 1 ea, plastic	60 ml minimum capacity with screw cap closure	60 ml with water tight closure, may be glass
Examining gloves 1 pr	size 76-108 (± 6) mm vinyl/ latex ambidextrous	Universal size, inert flexible elastomeric film or sheet or ambidextrous latex
Syringe, 1 ea	Filled with min. 9 ml of sterile water	30 ml filled with sterile water, USP
Cleaning Solution	22 to 29 ml	22 ml (min) in foil pouch, plastic or, glass container
Forceps, 1 ea	Required	Required
red rubber catheter outside diameter 5.0 cm (15 French)	N/A	N/A
Plastic tray, 1 ea, graduated in 100 ml increments	Minimum 1000 ml capacity 2 compartments	1000 ml capacity minimum, with 3 compartment lift out top
Absorbent cotton rayon balls	Sufficient size and quantity to prep patients	Minimum 5 ea 3 cm rayon, weight .91 grams or 3.81 cm cotton weight .3 grams each

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Component	Type III	Type IV
Underpad, 1 ea	30 x 30 cm minimum	46 x 41 cm ± 1.27 cm minimum
Fenestrated drape,	30 x 30 cm minimum	46 x 33 cm minimum, 9.5 cm approx. fenestration or 13.4 cm fenestration point to point diagonally
Lubricant, 1 ea	2 to 7 grams water soluble	5 grams, sterile water soluble
Specimen container 1 ea, plastic	30 ml minimum, with screw cap closure	60 ml minimum capacity with with screw cap closure, container may be glass
Examining gloves 1 pr	Size 95 (± 10) mm, plastic or latex ambidextrous	Size 76-108 (± 6) mm, latex, ambidextrous
Syringe, 1 ea	10 ml filled with either sterile water or saline	N/A
Cleaning Solution	10 ml, minimum	21 to 28 grams
Forceps	13 ± 1.3 cm long	13 ± 1.3 cm long
red rubber catheter outside diameter 5.0 cm (15 French)	N/A	Required

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Plastic tray. The tray shall contain all of the components packed in the order of use. The tray shall not leak when filled to capacity and shall be constructed to maintain its size and shape to the point of use. The minimum thickness of the bottom of the tray shall not be less than 0.0305 mm The base area of the tray with, respect to the total outside surface area, shall not be less than 22 percent. Type II tray shall be either compartmented or noncompartmented and shall be approximately 7.62 cm deep by 11 cm wide by 22 cm long If compartmented the main compartment shall have a fluid capacity of at least 700 ml Long side of tray shall have graduations which are visible from outside the tray.

For Type II kit, cleaning swabs (5 each) may be furnished in lieu of cotton/rayon balls, cleaning solution and forceps. Swab tip shall be approximately 3.81 cm in diameter and shall be premoistened with sufficient povidone-iodine solution to adequately prep patient Swabs or absorbent balls shall be provided in a suitable separate container.

Lubricant. Lubricant shall be provided in a suitable container marked or labeled with contents. Labeling shall meet the requirements of the Federal Food Drug and Cosmetic Act.

Specimen container. Shall be supplied with label for patient's name, room number, date, doctors name and instructions.

Examining gloves. Gloves shall be suitably packaged or folded to facilitate donning gloves without violating aseptic technique. The gloves shall be tested in accordance with ASTM designation D5151-90 or the lastest FDA glove adulteration watertight test procedure. Gloves shall be free from seams or welts in the palm or finger tip areas and shall conform to the human hand after not more than one minute wear. Refer to ASTM Designation D3578-91 for the metric conversion of glove sizes. Vinyl and plastic gloves shall be designated small, medium, large or universal size.

Cleaning solution. Cleaning solution shall be 10% Povidone-Iodine solution USP and shall be furnished in a suitable container with labeling meeting the requirements of the Federal Food, Drug and Cosmetic Act

Syringe For Type II, the syringe shall have a Luer taper tip with cap and shall be clearly and legibly graduated from 0 to 30 ml in 1 ml increments.

Catheter urethral. Should have an outside diameter of 5.0 cm (15 French), and shall be a Robinson red rubber catheter; sterile, 2 eyes, round tip, approximately forty one centimeters long.

Shelf life. Not less than 30 months (Type I and II) and 25 months (Type III and IV) of the expiration dating period shall remain at time of delivery to the government.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Material. The tray shall be fabricated from inert plastic and absorbent balls or swab tips shall be of natural cotton or viscose rayon. Underpad and fenestrated drape shall be absorbent and waterproof and fabricated from three-ply, laminated facial tissue or other suitable non-woven material. Underpad shall be coated on one side or backed with suitable plastic film to provide an impermeable barrier. Syringe and forceps shall be fabricated from suitable plastic.

Sterility. The components and internal surfaces of the package shall be sterile in accordance with the USP or AAMI requirements. Sterility assurance level shall be 0.000001.

Instructions. Each kit shall contain procedural instructions for use. In addition, all medical devices within the kit that contain natural rubber latex which comes directly or indirectly in contact with the body shall be identified on the principal display panel with the words ' THIS PRODUCT CONTAINS NATURAL LATEX RUBBER.' or where there is a concern about an allergic reaction, a similar statement, such as, "THIS PRODUCT CONTAINS NATURAL LATEX RUBBER WHICH MAY ALLERGIC REACTIONS IN SOME INDIVIDUALS '

Workmanship. The catheterization kit shall be free from defects which detract from its appearance or impair its serviceability. Components and tray shall be free from cuts, tears, holes, sharp edges, burrs, flash and foreign material.

Unit. KIT (KT). One catheterization kit packaged sterile, as specified, constitutes one unit. Components shall be packaged in tray in reverse sequence of use (i.e. first used components on top). For Type I kit the tray shall be CSR wrapped with a moisture resistant outer drape, minimum 48 x 51 cm, and overwrapped in protective plastic pouch, packaged to remain sterile until time of use. For Type II and III kits the individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened. For Type IV kits the tray shall be CSR wrapped 50.8 x 50.8 cm and inserted in a peel-open type polybag to remain sterile until use. Catheterization kit shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one catheterization kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Forceps performance test The forceps shall have no looseness or free play of either shank which would result in misalignment of the tips in closing. Upon closing, tips which do not meet shall be cause for rejection of forceps. This test shall be satisfactorily repeated a second time.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/ suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6515-00-519-3353	Catheterization Kit, Urethral	I
6515-00-117-9756	Catheterization Kit, Urethral Disposable, with 30 cc Syringe,	II
6515-00-334-3477	Catheterization Kit, Urethral, w/10 ml Syringe, Sterile, Disposable, 20's	III
6515-01-005-5483	Catherization Kit, Urethral, w/15 French, Red, Rubber Catheter	IV

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of United States Pharmacopedia (USP) regulations may be obtained from: USPC, Inc., 12601 Twinbrook Parkway, Rockville, MD 20852 USA.

Copies of ASTM Designation: D 3578-91, Standard Specification for Rubber Examination Gloves may be obtained from the American Society for Testing and Materials (ASTM), 1916 Race Street, Philadelphia, PA 19103-1187, USA

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order .

MILITARY INTERESTS.

Military Coordinating Activity:

DoD-MB

CIVIL AGENCY COORDINATING ACTIVITIES.

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

Project No. 6515-5532

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:36Z
Check the source to verify that this is the current version before use.