A-A-51455
20 May 1986

COMMERCIAL ITEM DESCRIPTION

TUBE, DRAINAGE, SURGICAL, THORACIC, Plastic, Radiopaque, Straight, Disposable, 28 Fr, 10s

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-T-37237A dated 28 February 1977.

This commercial item description covers one type of sterile, disposable, radiopaque, thoracic surgical drain tube and trocar suitable for closed intercostal drainage.

Salient characteristics:

General. The thoracic surgical drainage tube and trocar assembly shall consist of an outer plastic catheter and an inner aluminum trocar. The catheter shall be 28 French and the assembly shall be suitable for intercostal drainage.

Material.

Catheter. The catheter shall be fabricated of a transparent, clear, pliable, polyvinylchloride plastic meeting the United States Pharmacopeia (USP) requirements for Class IV plastics, for Transfusion and Infusion Assemblies. The lead content shall contain not more than 5 part per million (ppm).

Trocar. The trocar shall be fabricated of a suitable aluminum alloy.

Physical Properties.

Breaking strength. The minimum breaking strength of the catheter shall be 95 pounds.

Elongation. The minimum elongation of the catheter shall be 225 percent.

Dimensions. The overall length of the assembly shall be 17 $\pm$ 1/2 inches, and the catheter component shall be 16 $\pm$ 1/2 inches. The wall thickness of the catheter shall be 0.048 $\pm$ 0.005 inch.

FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

Style and design. The distal end of the catheter shall be uniformly chamfered to accommodate the trocar point, and the interior and exterior walls shall be smoothly finished. Two oval shaped eyes spaced 180° apart shall be provided at the distal end. One shall be approximately 3/4 inch from the tip, the other shall be approximately 1-1/2 inches from the first. A radiopaque stripe shall extend the entire length of the catheter. The aluminum trocar shall be fitted with a suitable ball type handles and it shall fit precisely into the catheter. The trocar point shall have a sharp conical tip with an included angle of 60 ± 5°.

Sterility. The catheter and trocar assembly shall be sterile when tested in accordance with the USP.

Workmanship. The item shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing ten catheter and trocar assemblies, each assembly individually packaged sterile in a container capable of maintaining sterility of the contents until package is damaged or opened, constitutes one unit.

Markings. Each individual assembly shall be permanently and legibly labeled with the manufacturers' name or registered trademark, French size, catheter length, commercial part number, and lot control number.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Inspection Level shall be S-2 with an AQL Level of 1.0 percent. The unit of product for test purposes shall be one catheter and trocar assembly.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-891-7593	TUBE, DRAINAGE, SURGICAL, THORACIC, Plastic, Radiopaque, Straignt, Disposable, 28 Fr, 10s

Ordering date (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

46447

Project No. 6515-3630

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.