A-A-51457A
30 January 1987

COMMERCIAL ITEM DESCRIPTION

HANDLE-ELECTRODE, ELECTROSURGERY

The General Services Administration has authorized the use of this commercial item description in preference to A-A-0051457(DM), dated 22 May 1986.

This commercial item description covers a sterile, disposable electrosurgical electrode handle for cutting and coagulation of tissues during surgical procedures.

Salient characteristics:

The unit shall be a sterile, disposable, pencil type electrosurgical handle, with a blade/needle electrode and a plastic safety cup holder.

The electrosurgical handle shall be fabricated of a suitable plastic material. The handle shall be fitted with a chuck that will accept and securely hold standard 3/32 inch (0.23 cm) shank diameter electrodes. The unit shall have a rocker type handswitch for cutting and coagulation activation.

The electrosurgical handle shall have a plastic safety cup holder which clips onto the sterile field. The holder shall have an electrode cleaning surface. The holder shall have a method for securing the electrode cord to prevent tangling of the cord and to maintain the cord within the sterile field.

The cord shall be approximately 10 feet(3 meters) long and shall be covered with a suitable insulation material. The cord shall be neither so stiff that it impedes manipulation, nor so flexible that it tangles easily. The cord shall have a formed male fitting to provide connection with an adapter (not included) fitting into the active outlet of an electrosurgical apparatus.

Workmanship. The electrosurgical handle shall be free from defects which detract from its appearance or impair its serviceability.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing twenty-five individually packaged sterile electrosurgical cautery handles, as specified, constitutes one unit. The individual package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-156-2794

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3963

Location: CID/ELECTRODE/H16-5

U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-038/70252

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.