A-A-51459
18 August 1986

COMMERCIAL ITEM DESCRIPTION

CLEANING COMPOUND, DENTAL TRAY

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a dental tray cleaning compound suitable for cleaning impression trays.

Salient characteristics:

The cleaning compound shall be a concentrate, powder type, which when mixed with water will produce a high strength liquid solution. Each bottle shall contain a minimum of 625 gm (1.4 lb.) of powder which when dissolved in water shall produce a minimum of 10 quarts (10.56 liters) of cleaning solution. The obtained solution shall be capable of effectively and rapidly cleaning alginate, zinc oxide - eugenol, modeling compounds and dental waxes from metal impression trays. The cleaning compound shall be soluble in hot, warm or cold water. The cleaning compound shall be harmless to skin, non-toxic, non-corrosive, and non-irritating, when used for the purpose intended.

Material. The cleaning compound shall be an alkaline surfactant that forms an anionic solution comprised of wetting agents, emulsifying agents and carbonates. The cleaning compound shall not contain phosphates. The solution when mixed in accordance with manufacturer's instructions shall have a pH value of 9.0 to 9.9.

Instructions for use shall be provided with each unit for issue.

Workmanship. Powder shall be free of lumps, dirt and foreign matter and from other defects which detract from its appearance or may impair its serviceability.

Unit. Bottle (BT). Six hundred twenty-five (625) grams of powder in an unbreakable, resealable bottle, with instructions, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one bottle of tray cleaning compound. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packing, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment, from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6520-01-145-6339	CLEANING COMPOUND, DENTAL TRAY Detergent Concentrate, 625 Grams, Powder.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activity:
VA-OSS

51249

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Project No. 6520-1902

Location: NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.