A-A-51463
6 June 1986

COMMERCIAL ITEM DESCRIPTION

SUCTION SET, TRACHEAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers two sizes of disposable, sterile tracheobronchial suction sets.

Salient characteristics:

Sizes. Sets shall be provided, as specified, with catheters of the following sizes:

Size	Catheter outside diameter-(French)
1	14 ± 1
2	18 ± 1

General. The set shall consist of one catheter, one solution tray, and one glove.

Catheter. The catheter shall consist of transparent flexible tubing 51 plus or minus (±) 2 centimeters (cm) in length attached to a tubing adapter. The tubing shall have a whistle tip on the distal end. The opening of the whistle tip shall be 0.25 to 0.35 cm in diameter. The opening shall have smooth edges. In addition, the distal end of the tubing shall have one or two side eyes. The side eye shall be within 1 cm of the distal tip, and shall have a cross sectional area of 0.6 to 0.9 cm. The tubing adapter shall have a thumb control port raised 0.5 to 1.0 cm above the surface of the adapter. The port shall consist of a flange 2.0 ± 0.5 cm by 1.5 ± 0.5 cm with an elongated opening to the adapter lumen in the center of the flange. The cross-sectional area of the opening shall be 1.5 to 2.2 times that of the adapter lumen. The proximal end of the tubing adapter shall make a secure connection with 1/4 to 5/16 inch inside diameter plastic tubing.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Solution tray. The tray shall be a disposable, collapsible container. The tray shall be capable of holding fluids for one hour without becoming saturated with liquid or leaking. The tray shall hold a minimum of 240 milliliters of solution.

Glove. The glove shall be pre-cuffed. The glove shall be size medium or large and shall meet the requirements of American Society for Testing and Materials (ASTM) Standard D3578.

Materials. The catheter tubing shall be manufactured from flexible polyvinychloride (PVC). The catheter tubing adapter shall be manufactured from a rigid plastic. The glove shall be manufactured from vinyl or latex.

Physical characteristics. PVC material shall be tested for tensile strength and ultimate elongation in accordance with ASTM Standard D638. The catheter tubing material shall have a minimum tensile strength of 1500 pounds per square inch (psi) and a minimum ultimate elongation of 200 percent. The tubing material shall be tested for hardness in accordance with ASTM D2240, and shall have a Shore Durometer hardness of A77-A80. A sufficient amount of material shall be tested as to be representative of the lot. Any test failure shall be cause for rejection.

Sterility. The components of the set shall be sterile when tested in accordance with the United States Pharmacopeia (USP).

Toxicity. PVC used in the catheter tubing shall meet the requirements of the USP for Class IV plastic.

Workmanship. The finished catheter shall be free from loose flash cracks, burrs, or sharp edges. There shall be no defects which affect the life, appearance, or serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). Each catheter set shall be individually packaged in a peel-open type container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 100 catheter sets of one size only, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Suction sets shall be examined to determine compliance with all requirements contained in this specification.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection level II, AQL 2.5 (% defective). The unit of product for examination shall be one suction set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-123-1731	CATHETER SET, SUCTION, STERILE, Tracheobronchial, Disposable, 18 Fr, 100's
6515-01-167-6666	CATHETER SET, SUCTION, STERILE, Tracheobronchial, Disposable, 14 Fr, 100's

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS

PGC 06647

Project No. 6515-3586

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.