A-A-51464
9 June 1986

COMMERCIAL ITEM DESCRIPTION

RESERVOIR, SUCTION, JACKSON-PRATT, SILICONE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile disposable suction reservoir for use with Jackson-Pratt style surgical drains.

Salient characteristics:

Material. The reservoir shall be manufactured from a transparent silicone rubber compound. The reservoir material shall be sufficiently transparent to see graduations when viewed from the opposite side of the finished reservoir.

Physical properties. The silicone compound used in the manufacture of the bulb proper shall have the properties shown in Table I. Silicone compound shall be tested for the property using the American Society for Testing and Materials (ASTM) Standard shown in Table I.

Table I. Physical Properties and Test Methods

Property	Value	ASTM Standard
hardness (Shore durometer)	50-70A	D-2240
tear resistance	200 pounds force per inch (minimum)	D-624
tensile strength	950 pounds force per square inch (minimum)	D-412
elongation	400 percent (minimum)	D-412

Style and design. The reservoir shall be an oval-shaped aspirator bulb with a permanently attached semi-rigid silicone rubber head. The reservoir head shall have a drainage port, a suction port, and an attached bulb strap. The suction port shall have a plug attached to it which makes a mechanically secure and leak-proof connection with the drainage port.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

The suction port shall protrude from the reservoir head 0.500 ± 0.125 inches. The suction port shall have three steps ranging in outside diameter from 0.156 ± 0.007 inches at the top of a step to 0.187 ± 0.007 inches at the base of a step. Each step shall be 0.140 ± 0.007 inches in length, with the top of the first step forming the distal tip of the suction port. The drainage port plug shall be attached to the suction port between the third step base and the reservoir top. The suction port shall have an inside diameter of 0.110 ± 0.007 inches. The suction port shall connect to an integral Bunsen valve on the interior of the reservoir which allows flow of fluids into the reservoir but prevents flow from the reservoir into the suction port. The Bunsen valve shall also prevent the loss of negative pressure in the suction tubing when the reservoir drainage port is opened.

The reservoir drainage port shall protrude from the reservoir head 0.250 ± 0.020 inches, shall have an inside diameter of 0.156 ± 0.010 inches, and shall have an outside diameter of 0.219 ± 0.01 inches.

The reservoir shall have an overall length of 4.75 ± 0.25 inches when measured from the suction reservoir tip to the bottom of the reservoir. The reservoir shall have an outside diameter of 2.20 ± 0.2 inches at the point of maximum diameter. The reservoir shall have a minimum internal volume of 125 milliliters (ml).

Graduations. The reservoir shall be graduated in 25 ml increments from 25 to 100 ml. Each graduation line shall extend around the circumference of the bulb and shall be numbered. The reservoir shall bear the letters "cc" or "ml" beside or near the 100 ml graduation. Graduations shall be accurate to within ± 10 percent of the labeled volume. Accuracy shall be tested with water and a graduated cylinder.

Performance. Finished reservoirs shall be tested in accordance with the vacuum and flow rate test.

Vacuum test. Attach a manometer or pressure gage to the suction port. Manually or mechanically evacuate the air from the reservoir and install the reservoir drain plug. Release the reservoir. The reservoir shall create a minimum vacuum of 85 mm of mercury, and shall hold the vacuum for a minimum of 2 minutes.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

Flow rate test. Mount the reservoir by the strap above a supply basin of water at room temperature such that a hydrostatic head of 7 + 1/8 inches exists from the water level in the basin to the entrance of the reservoir. The reservoir may be held by a mechanism other than the strap as long as the reservoir shape recovery and the vacuum induced are not influenced by the mechanism and the 7 inch head is maintained. Manually or mechanically evacuate the air from the reservoir, install the reservoir drain plug, and release the reservoir. The reservoir shall collect 33-45 ml of water in 2 minutes.

Sterility and Pyrogenicity. Tubes shall be sterile and free from pyrogens. Tubes shall be tested for sterility and pyrogenicity in accordance with the United Stated Pharmacopiea (USP).

Biocompatibility. Material used in the manufacture of the reservoir shall be tested in accordance with the USP Muscle Implantation Test and USP Intradermal Sensitivity Test.

Workmanship. The item shall be processed in such a manner as to be uniform and free from defects that will affect life, serviceability, or appearance.

Unit. Package (PG). One package containing ten sterile reservoirs, each reservoir individually packaged in a suitable peel-open type container capable of maintaining sterility of contents unless package is damaged or opened, as specified, constitutes one unit.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Reservoirs shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one reservoir. Sampling shall be S-2, AQL 2.5

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-216-3369	RESERVOIR, SUCTION, JACKSON-PRATT, Silicone

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS

PGC SF023

Project No 6515-3697

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.