A-A-51465
16 June 1986

COMMERCIAL ITEM DESCRIPTION

CONTAINER, URINE SPECIMEN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a collapsible container suitable for the collection and storage of urine.

Salient characteristics:

Materials. The urine specimen container shall be fabricated of a durable, high quality plastic coated chipboard containing an inner liner of a strong, high density polyvinyl. The inner liner shall be clear plastic allowing ready view of contents.

Coating. The plastic coating shall be a non-toxic thermoplastic (polyethylene) or a copolymer of vinyl chloride, vinyl acetate or a plastic coating which shall be free of any disagreeable color and shall not contaminate or react with the intended end use contents (chemically inert).

Ink. The ink used for markings shall be waterproof ink generally used by the food packaging industry.

Design and construction. The urine specimen container shall consist of a plastic bag located within a coated paper container. The plastic bag shall be secured to the container at the opening through which urine will pass. The container shall be pre-folded for convenient storage. The container shall be collapsible and shall be capable of being easily assembled for use. The container shall be capable of holding a minimum of 2500 cc of urine; and shall employ some means for approximating the measurement of volume in 100 cc increments such as viewing windows. The container shall have an integral carrying handle; and shall provide means for reclosing the top after each use. The container shall withstand refrigeration without degradation of materials or loss of functional capabilities. The container shall comply with all applicable FDA regulations. The container shall have a suitable area on the container for writing the patient's name and date.

Closure. The container shall consist of a device such as a plastic clip for sealing the spout and thus preventing spillage during transport.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Capacity. The container capacity shall be at least 2500 cc.

Graduation markings. Graduations shall be in 100 cc steps beginning at not less than 300 cc and ending at not less than 2500 cc. The following graduation markings shall at least be identified: 500 cc, 1000 cc, 1500 cc, 2000 cc, and 2500 cc. All letters shall be not less than 1/8 inch in height.

Instructions. Each individual unit shall have information printed on the outside of the container or on a separate sheet or paper regarding proper use of the container.

Workmanship. The urine specimen container shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one hundred specimen containers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one urine specimen container.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-998-9478

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

i-A-51465

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6530-1836

Location: CID/CNTR-URINE-SPEC/H10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.