A-A-51468
10 June 1986

COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET, With Flexible Drip Chamber, Non-Vented, Adult, Without Needle, Disposable, 50's

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a single type of non-vented intravenous injection set used in administering parenteral solutions from glass containers with airways or flexible bags.

Salient characteristics:

Materials. The tubing shall be clear, plasticized polyvinyl chloride having a Shore Durometer A hardness of 78 ± 5. Rubber at injection sites shall be of the best quality and suitable in all respects for injection and resealing after needle withdrawal. The clear polyvinyl chloride sidewall of the drip chamber shall have a Shore Durometer A hardness of 96 ± 5, -10. All plastic and rubber components shall be sterile, pyrogen-free and, if required, to be in contact with the parenteral fluid, in compliance with the USP requirements for Class IV plastics. The spike and cap of the drip chamber, the roller clamp, the Y's, and male Luer taper shall be made of suitable rigid plastic.

Construction. The set shall be non-vented and include a spike, flexible drip chamber calibrated at 15 ± 3 drops/ml, a check valve, two Y-injection sites, a roller clamp to provide a constant flow setting and stoppage, a flash tube and male slip fit Luer taper complying with ANSI/HIMA MD70.1. The overall length of the set shall be 90 ± 5 inches (229 ± 13 cm). The drip chamber wall shall be 0.019 to 0.030 inch thick, have an ID of 0.460 to 0.490 inch, and be about 2-3/4 inches high. The check valve shall be about 14 inches below the drip chamber, the upper Y injection site shall be about 20 inches below the drip chamber, the lower Y injection site shall be about 8 inches from the Luer tip, and the roller clamp shall be between the two Y injection sites. The tubing shall have a 0.019 to 0.025 inch wall and be 0.106 ± .005 inch in ID. All fittings shall be permanently attached and leakproof.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

The IV set shall withstand an internal hydraulic pressure of 20 psig for five minutes without leakage of no more than one falling drop.

The roller shall not be able to be forced out of the clamp body under a 3.65 kg load applied axially and toward the open end of the clamp. End protectors shall be present to maintain sterility. The set shall be packaged in a suitable container to maintain its sterility until the package is opened.

The two Y injection sites and the flash tube shall reseal without a falling drop after puncture and withdrawal of a non-lubricated 18-gage needle. the injection sites, within 10 seconds of needle withdrawal, shall be subjected to an internal pressure of 4 psig for 30 seconds.

A. Sterility test. The catheter and needle unit shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Workmanship. The set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 50 intravenous injection sets, each set individually packaged sterile in a sealed, peel-open container, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one intravenous injection set.

Sampling Criteria for Tests

	Inspection level	AQL
Pressure test	S-3	0.65
Security of roller in clamp body	S-3	0.65
Resealing of injection sites	S-3	0.65

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-105-0614	INTRAVENOUS INJECTION SET, With Flexible Drip Chamber, Non-Vented, Adult, Without Needle, Disposable, 50s

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order.)

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:56Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51468

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6515-3567

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.