A-A-51471
10 June 1986

COMMERCIAL ITEM DESCRIPTION

MALLET, NASAL, COPPER, Small

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a small copper, nasal, bone mallet suitable for use in nasal surgery.

Salient characteristics:

General. Shall be a small nasal bone mallet suitable for use during nasal surgery.

Material. The mallet head shall be fabricated from copper rod or shall be a stainless steel jacket with a copper insert or as a minimum shall have a copper striking surface. The mallet head shall not be plated. The mallet shaft shall be fabricated from 410 or 420 stainless steel in accordance with ASTM F899. Should 410 stainless steel be used, the steel shall have a Rockwell hardness of C40 to C48. Should 420 stainless steel be used, the steel shall have a Rockwell hardness of C45 to C52. The handle shall be fabricated from Class 3 austenitic stainless steel in accordance with ASTM F899. The handle shall have a minimum Rockwell hardness of B75. The mallet shaft and handle shall be tested for Rockwell hardness in accordance with ASTM E18.

Characteristics. The nasal mallet shall have a total weight of approximately 4 ounces. The nasal mallet shall have an overall length of 6.5 inches plus or minus 0.5 inch provided the design achieves an ideal balance. The mallet head shall have a diameter of 0.75 inch plus or minus 0.063 inch and an overall length of 1.25 inches plus or minus 0.063 inch. The handle design shall provide a secure grip. The handles may be fluted or knurled.

Marking. The mallet shall be marked with the name or registered trademark of the contractor. The word "stainless" shall be marked on the shaft. If the mallet is of foreign manufacture, the mallet shall be marked with the country of origin.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.

Boil test. The nasal mallet shall meet the boil test requirements as specified in GG-I-526.

Finish. The finish on all edges and surfaces shall be smooth, free of burrs, sharp edges, cracks, crevices or rough areas.

Workmanship. The mallet shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One mallet, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. AQL shall be 1.0 percent defective. Inspection level shall be S-2. The unit of product for test purposes shall be one mallet.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-141-3330	MALLET, NASAL, COPPER, Small

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION A-A-51471

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC - S0011

Project No. 6515-3684

U.S. GOVERNMENT PRINTING OFFICE 1988 — 605 036/43406

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.