A-A-51473
16 June 1986

COMMERCIAL ITEM DESCRIPTION

BASIN, SITZ BATH

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable plastic sitz bath designed for single patient use in medical treatment facilities or at home.

Salient characteristics:

Shall be a plastic sitz bath that fits easily on a toilet or commode and is for single patient use without the aid of a medical assistant.

The plastic bowl shall be high impact polyethylene and shall be suitable for perineal and rectal care.

A flexible plastic bag suitable for suspending and for holding at least 1900 ml of water shall be provided. The bag shall have a water tight filler mechanism at the upper end of the bag.

Attached to the bag shall be 60 inches of tubing that shall allow a steady flow of water and that shall have an easy means of restricting and stopping the flow.

The bowl shall have a working width of at least 9 1/2 inches (24.1 cm), shall have an entrance for the tubing into the bowl, and a device for holding the tubing in the bottom of the bowl. The bowl shall be contoured to sit directly on a toilet bowl and shall have generous rear openings for front-to-rear overflow.

The plastic bowl shall be free from objectionable odor, sharp edges, cracks, holes or thin areas. The edges shall be smooth or rounded to preclude injury to patients.

Each individual sitz bath with associated bag and tubing shall be enclosed in an opaque plastic bag.

Instructions for use shall be provided with each unit of issue.

Workmanship. The sitz bath shall be free from defects which detract from its appearance or impair its serviceability.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing ten sitz baths, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one sitz bath.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:56Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-030-6861

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS

PGC 50626

Project No. 6530-1833

Location: CID/BASIN-SITZ-BATH-H13

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44941

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:56Z
Check the source to verify that this is the current version before use.